



## CERTIFICATE OF CALIBRATION

ADL/BO/MS/C/0311/22-23

**Customer Name** : Radvision Diagnostics & Labs

**Address of the customer** : Flat No: 1, Ground Floor, Reshma Apartments, Nandidurga Road, Jayamahal, Bangalore

**Instrument Model** : Mispa Count X - 3 Part Hematology Analyzer

**Serial No.** : 2121100010194

**Date of Installation** : 21/09/2021

**Date of Calibration** : 25/Jan/2023

This is to certify that the above mentioned product has been calibrated for CBC parameters (WBC, RBC, HGB, HCT, PLT) according to the standard operating procedure given by Agappe

The reference instruments used for value assignment are managed by the traceability system in

*Count 2 Series Instrument* and these are traceable to the International Standards such as - CLSI & ICSH

For AGAPPE DIAGNOSTICS LTD.

### Authorized Signatory



Name: Femim Simon  
Designation: Engineer Customer Service

**AGAPPE DIAGNOSTICS LTD.** ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC335413

**DELHI OFFICE / REAGENT PLANT**  
20, Patilnagar (PO), Dist. Ernakulam, Kerala - 683 562, India.  
Ph: 0475 7900 | Email: agappe@agappe.in

**EQUIPMENT PLANT**  
X/500-CB, Block No. 32, KJWRA Small Industrial Park,  
Nellai, Cochin, Kerala, India - 686 721. Tel: +91 484 274 7477.

**KOLKATA OFFICE**  
406, North Midra Plot No-10, Block-DK, Sector-V,  
Salt Lake City, Kolkata - 700 095.  
Tel: +91 20 4099 0850 | Email: kolkata@india.agappe.in

**REGISTERED OFFICE**  
4, 6th Floor, Anandh Business Centre, 179,  
S. Jeevanandhi, Andheri (East), Mumbai - 400 099, India.

**DELHI OFFICE**  
D540 540, 5<sup>th</sup> Floor, DLF Phase, Gurgaon Marg,  
New Delhi - 110 001, India.  
Tel: +91 11 4210 8112 | Email: delhi@india.agappe.in

**BANGALORE OFFICE**  
5-6, 6<sup>th</sup> Floor, Red Cross Bldg, No.20, Salt Lake Road,  
Bangalore - 560 002. Tel: +91 80 2228 8208  
E-mail: bangalore@india.agappe.in

DETAILS OF CONTROL RUN

Instrument Model : Mispa Count X - 3 Part Hematology Analyzer  
 Serial No : 2121100010194  
 Date of Control Run : 2023 - 02 - 14  
 Control Used : Hematology Control (Tri Level)  
 Lot No : 32110386  
 Expiry : 2023 - 03 - 05

The results of repeatability test are summarized in the following table. Refer the attached control sheets for details.

Sample No	WBC K/ $\mu$ L	RBC M/ $\mu$ L	HGB g/dL	HCT (%)	PLT K/ $\mu$ L
1	7.1	4.3	12.7	36.5	268
2	7.0	4.4	13.0	38.3	272
3	7.1	4.2	13.2	38.	265
Instrument Mean	7.0	4.3	12.9	37.6	268
Range	6.2 - 7.8	3.5 - 5.3	12.1 - 13.9	32.5 - 43.5	231 - 295
CV %	11.42	20.9	6.97	14.6	11.94

This is to certify that the above mentioned product has been calibrated for CBC parameters (WBC, RBC, HGB, HCT, PLT) according to the standard operating procedure given by Agappe.

The control run after calibration shows good repeatability and precision

Next Calibration due on 24/July/2023

For AGAPPE DIAGNOSTICS LTD.

Authorized Signatory



Name: Femin Simon

Designation: Engineer Customer Service

**AGAPPE DIAGNOSTICS LTD.** ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

**CORPORATE OFFICE / REAGENT PLANT**  
 Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.  
 Tel: + 91 484 286 7000 | Email: agappe@agappe.in

**EQUIPMENT PLANT**  
 X/588-CB, Block No. 32, KINFRA Small Industrial Park,  
 Nettad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

**KOLKATA OFFICE**  
 406, Merin Matrix, Plot No-10, Block-DN, Sector V,  
 Salt Lake City, Kolkata - 700 091.  
 Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

**MUMBAI (REGISTERED OFFICE)**  
 401 & 402, 4<sup>th</sup> Floor, Jaisingh Business Centre, 119,  
 Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.  
 Tel: +91 22 4200 0000 | Email: mumbaioffice@agappe.in

**DELHI OFFICE**  
 DSM 540, 5<sup>th</sup> Floor, DLF Tower, Shivaji Marg,  
 New Delhi - 110 015, India.  
 Tel: +91 11 4000 0416 | Email: delhioffice@agappe.in

**BANGALORE OFFICE**  
 5-6, B<sup>th</sup> Floor, Red Cross Bhawan, No.26 Race Course Road,  
 Bangalore - 560 001. Tel: +91 80 2226 8288  
 Email: bangaloreoffice@agappe.in