

Calibration Certificate

This is to certify that Automated Hematology Analyzer XN-550 Serial No: 20207 installed at District Hospital Sehorehas been calibrated.

Calibration Done on 5thMarch 2022

Calibration Validity: 5th March 2022 to 4thMarch 2023

Calibration Due on 4thMarch2023

WHOLE BLOOD CALIBRATION DETAIL

LOT NO.- 20522101 VALID UP TO- 05/03/2022 EXPIRY- 04/03/2023

Parameter	WBC	RBC	HGB	HCT	PLT
1	7.03	4.38	12.2	35.1	238
2	7.06	4.38	12.2	35.1	244
3	7.10	4.32	12.1	34.7	237
4	6.96	4.32	12.1	34.8	251
5	7.06	4.36	12.2	35.0	250
6	7.13	4.36	12.2	35.0	242
7	7.10	4.32	12.2	34.8	238
8	7.16	4.36	12.2	34.9	248
9	7.00	4.33	12.2	34.7	246
10	6.96	4.37	12.2	35.1	248

11	6.96	4.35	12.2	35.1	247	2.28	4.13	265
MEAN	7.049	4.347	12.18	34.91	245.1	2.373	4.185	269.8
TARGET	7.088	4.407	12.19	35.30	252.3	2.451	4.259	260.8
CV %	1.1	0.5	0.3	0.4	2.0	3.2	0.8	1.3
Acceptable Limit %	3.0	1.5	1.0	1.5	4.0	15.0	2.5	6.0

Calibration accepted for above mentioned values

For Sysmex India Pvt. Ltd.

Authorized Signatory

Sysmex India Pvt Ltd (CIN U33120MH1998PTCT3943)
Office 1002, Damji Shamji Business Galleria, 10th Floor, LBS Marg, Kanjur Marg (West), Mumbai - 400078, Maharashtra, India Tel. +91 (22) 6112 6666
Factory Village Malpur, Pargana Dharampur, Nalagarh Road, Baddi - 173205, Dist. Solan, Himachal Pradesh, India Tel. +91-9218422282 / 9816672282
www.sysmex.co.in