

Date: 28thJune2023

Calibration Certificate

This is to certify that Automated Hematology Analyzer XN-550 Serial No: 20207 installed at District Hospital Sehore has been calibrated.

Calibration Done on 28th June 2023

Calibration Validity: 28th June 2023 to 27th June 2024

Calibration Due on 27th June 2024

WHOLE BLOOD CALIBRATION DETAIL

LOT NO.- 31632101 VALID UP TO- 25/06/2023 EXPIRY- 16/07/2023

Parameter	WBC	RBC	HGB	HCT	PLT
1	6.90	4.31	11.9	34.3	264
2	6.90	4.30	11.9	34.3	263
3	6.91	4.26	11.8	34.2	257
4	6.98	4.27	11.9	34.2	250
5	6.88	4.27	11.9	34.3	248
6	6.88	4.29	11.9	34.4	250
7	6.76	4.24	11.8	34.0	243
8	7.06	4.29	11.9	34.4	252
9	6.78	4.26	11.9	34.3	248
10	6.97	4.29	11.8	34.4	256

11	6.89	4.29	11.9	34.3	245
MEAN	6.901	4.277	11.87	34.28	251.1
TARGET	7.056	4.348	11.94	34.74	247.9
CV %	1.3	0.4	0.4	0.4	2.4
Acceptable Limit %	3.0	1.5	1.0	1.5	4.0

Calibration accepted for above mentioned values.

For Sysmex India Pvt. Ltd.

Authorized Signatory

Sysmex India Pvt Ltd 1001 123000410000010145
 Office: 1002, Durg Shamji Business Galleria, 10th Floor, LBS Marg, Colaba (West), Mumbai - 400075, Maharashtra, India. Tel: +91 (22) 6012 5444
 Factory: Village Malpuri, Patana Dharampur, Halagani Road, Bhubaneswar - 751005, Dist. Sakh, Orissa, India. Tel: +91 6742611100 1501111100
 www.sysmex.co.in