

BIO-RAD LABORATORIES (INDIA) PRIVATE LIMITED

Site name:	District Hospital Sehore
Address:	Sehore (M.P)
Province-Postal code	466001
Contact Name	Mr. Pradeep Chandrawanshi

Calibration Certificate

1	08/07/2022	Instrument Model	D-10 Hb Analyzer	Instrument	DM0G525924
-	(dd mmmm, yyyy)	Model	5 10 110 7 mary 201	Serial Number	DIVIOG323924

We certify that the Bio-Rad instrument specified above and installed in your lab is calibrated and the performance was found satisfactory.

High Pressure Pump Flow RateReference: Service Manual: Section 7.7.3

Measured with (Device manufacturer and model	Device Serial Number	Device Calibration date (dd mmmm, yyyy)
Measuring cylinder hex base 10ML	10115	10/06/2022

Specified Range@ 1.5ml/Minute	Measured Value
1.425 - 1.575 ml/minute	1.492

Column Holder temperature Reference: Service Manual: Section 7.9

Measured with (Device manufacturer and model	Device Serial Number	Device Calibration date (dd mmmm, yyyy)
FLUKE 179	2026569	27/04/2022
Specified Range @ 35°C	Measured Value	•
33.4°C – 36.4°C	34.5	

Detector log value Reference: Service

Reference: Service Manual: Section 7.2 or Field Bulletin D-10-081

of a large a definition of the second of the	1101a Danctii D-10-001		
Measured with	Specified Range	Measured Value	
Instrument Service Software	0.010 0.022 411	0.021AU	

The next calibration is due on	08/07/2023	(dd mmmm, yyyy)
		(aa mymmin, yyyy)

For Bio-Rad Laboratories India Pvt Limited

Authorized Signatory:

Page 1 of 1

Calibration Certificate