

ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



CALIBRATION CERTIFICATE

Customer name	M/S: District Govt. Hospital	Service request No. & date	P/03/12/06/23&12/06/2023
And	Sehore	ULR no.	CC233923000003212F
Address	Central Processing Lab Sewan	Cert. No.	AT23000003212
	River Square Sehore Mp 466001.	Date of Receipt of DUC	12/06/2023
		Date of calibration	12/06/2023
		Date of issue	13/06/2023
		Suggested due date	11/06/2024

	Instrument De	etails	
Instrument name	Centrifuge m/c	Location	
Make / Model No.	Remi/R-8C	Accuracy	
Range / Size	0 to 5000 rpm	Visual Inspection	OK
Least Count	1 rpm	Calibration Performed at	Site
I.D. No. / Sr. no.	She/Cpl/Centrifuge/13/ Hhlc- 12207		

Detail of reference standards & Major equipments used			
Equipment Name	Digital Tachometer		
Make	LT Lutron		
Model / SR No.	/A.C92792		
Certificate No.	QCTS/000246/10		
Calibration Validity	20/10/2023		
Calibration by	Quality Calibration Testing Sol.		

Environmental	Temperature	(20±2)°C	Calibration Reference	IS: 12508 (1988)
Condition	Relative Humidity	(50±10) %	Work Instruction	AT/WI-31M

Serial No.	Standard value in (rpm)	Measured value in (rpm)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	503.7	502 ·	
02.	1002	1000	±0.20 %rdg
03.	2011	2009	
04.	3005	3003	
05.	3515	3513	
06.	4009	4005	

Remarks:

(1) Standard equipment use for ealibration are traceable to national/ international standards. ---

(2)The reported expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor k = 2 such that the coverage probability corresponds to approximately 95%.(3) The above results are valid at the time of and under the stated conditions measurement (4) This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By (Calibration Engg./TM) (MAHIPAL) Form No.- QF-47



Approved By (QM/TM) (EERAJ TYAGI) Page No. 1 of 1