



Quality Services & Laboratories

Plot No.10, Second Floor, D.S.I.D.C. Scheme - III
Okhla Industrial Area, Phase - II, New Delhi- 110020, India,
Tel. : +91 85953 43496
Email: corp@qsl.co.in ; Web.: www.qslglobal.com



CC-2800

CALIBRATION CERTIFICATE

Unique Lab Report No. : ULR-CC290023000004604F
Certificate No. : QSL/DEL/CAL/23061205.03
Date of Calibration: 12.06.2023
Date of Receipt of Instruments: 12.06.2023

Page 1 of 1
SRF No. : QSL/23061205
Suggested Due Date: 12.06.2024
Date of Issue: 14.06.2023

Customer Name & Address:

M/s: DISTRICT GOVT. HOSPITAL SEHORE
CENTRAL PROCESSING LAB SEWAN RIVER SQUARE
SEHORE (M.P) 466001

Description of Instrument:

Name :	Micro Pipette	Range :	5 – 50 μ l
Make :	Dragon	Least Count :	0.1 μ l
S. No.:	YE202AM0109739	Accuracy :	—
Model No.:	—	Performed at :	Lab
Location:	Central Processing Lab	Visual Inspection :	OK
Customer ID:	SHE/CPL/Dragon/09	Zero Error :	Nil

Environmental Condition:

Temperature :	25°C \pm 3 °C	Humidity :	40% to 60%
---------------	-----------------	------------	------------

Methodology of Calibration:

Reference Standard: IS: 4787	Work Instruction No. :	QSL/WI/M/21
------------------------------	------------------------	-------------

Standard & Major Equipment (s) Used for calibration:

S.No.	Instrument Name	Make	S. No./Model No.	Certificate No.	Calibrated By	Valid Upto
1.	Std.Weight Box	Weightronics	—	WMCLIF/2023-06/2469	Weightronics	08.06.2025

Calibration Result:

UUC Reading in μ l	Standard Reading in μ l	Error in μ l	Uncertainty of Measurement
10	9.80257	0.19743	\pm 0.9 μ l
20	19.79521	0.20479	
50	49.67036	0.32964	

(The reported expanded Uncertainty in measurement is stated as the standard uncertainty in measurement multiplied by the coverage factor $k = 2$, which for a normal distribution corresponds to a coverage probability of approximately 95%)

Note:

- *The calibration results reported in this calibration certificate are valid at the time of & under stated conditions of measurement.
- *This certificate cannot be reproduced except in full without our prior permission in writing.
- *This certificate refers only to the particular item (s) calibrated.
- *Laboratory standard are traceable to national standards.
- *UUC: Unit Under Calibration

Prepared By

Customer Service Cell (Gajendra Singh)



Approved By

CEO (Mridul Kohli)

OSL-1Q-127

All business are undertaken subject to company standard general business conditions, which are available on request.