



# Quality Services & Laboratories

Plot No.10, Second Floor, D.S.I.D.C. Scheme - III  
Okhla Industrial Area, Phase - II, New Delhi- 110020, India,  
Tel. : +91 85953 43496  
Email: corp@qsl.co.in ; Web.: www.qslglobal.com



## CALIBRATION CERTIFICATE

Unique Lab Report No. : ULR-CC290023000004605F  
Certificate No. : QSL/DEL/CAL/23061205.04  
Date of Calibration: 12.06.2023  
Date of Receipt of Instruments: 12.06.2023

Page 1 of 1  
SRF No. : QSL/23061205  
Suggested Due Date: 12.06.2024  
Date of Issue: 14.06.2023

### Customer Name & Address:

M/s: DISTRICT GOVT. HOSPITAL SEHORE  
CENTRAL PROCESSING LAB SEWAN RIVER SQUARE  
SEHORE (M.P) 466001

### Description of Instrument:

Name :	Micro Pipette	Range :	100-1000 µl
Make :	Dragon	Least Count :	1 µl
S. No. :	YE202AM0109739	Accuracy :	---
Model No. :	----	Performed at :	Lab
Location :	Central Processing Lab	Visual Inspection :	OK
Customer ID :	SHE/CPL/Dragon/10	Zero Error :	Nil

### Environmental Condition:

Temperature :	25°C ± 3 °C	Humidity :	40% to 60%
---------------	-------------	------------	------------

### Methodology of Calibration:

Reference Standard: IS: 4787	Work Instruction No. :	QSL/WI/M/21
------------------------------	------------------------	-------------

### Standard & Major Equipment (s) Used for calibration:

S.No.	Instrument Name	Make	S. No./Model No.	Certificate No.	Calibrated By	Valid Upto
1.	Std.Weight Box	Weightronics	----	WMCL/F/2023-06/2469	Weightronics	08.06.2025

### Calibration Result:

UUC Reading in µl	Standard Reading in µl	Error in µl	Uncertainty of Measurement
100	99.73320	0.26680	± 0.9 µl
500	499.60897	0.39103	
1000	999.50279	0.50279	

(The reported expanded Uncertainty in measurement is stated as the standard uncertainty in measurement multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95%)

### Note:

- \*The calibration results reported in this calibration certificate are valid at the time of & under stated conditions of measurement.
- \*This certificate cannot be reproduced except in full without our prior permission in writing.
- \*This certificate refers only to the particular item (s) calibrated.
- \*Laboratory standard are traceable to national standards.
- \*UUC: Unit Under Calibration

Prepared By

Customer Service Cell (Gajendra Singh)



Approved By

CEO (Mridul Kohli)

QSL -1Q-129