



Abbott (Diagnostics Division)

405, KIRTI SHIKHAR BUILDING,
JANAKPURI DIST. CENTRE,
NEW DELHI - 110 058
TEL. +91 11 25619194 FAX. +91 11 25613392

Dated-April 7th, 2023

Certificate Of Calibration

This is to certify that the below mentioned instrument installed in Redcliffe Lifetech Pvt Ltd, Jodhpur calibrated by us on 7th April, 2023. The next calibration is due on or before 6th April, 2024. The values are falling within the specified range.

Instrument Model- Ci4100
Serial No. -C403143/I1sr65500
Software Version - 9.50


For Abbott Healthcare Pvt Ltd

Prashant Mishra
Technical Service Specialist
Abbott Diagnostic Division

QC Results List Report

Released

Operator ID: ADMIN

System serial number: 86357

C / P	Module	SID	Control name	Assay	Result	Flags	Code	Date / Time
M906/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	AlbBCG2	3.9 g/dL			11.07.2023 10:55
M910/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	AlkP	99 U/L			11.07.2023 10:23
M910/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	ALT	26 U/L			11.07.2023 10:26
M910/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	AST	39 U/L			11.07.2023 10:25
M910/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	BilD	0.47 mg/dL			11.07.2023 10:25
M910/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	BiliT	1.10 mg/dL			11.07.2023 10:22
M910/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	CaC	9.65 mg/dL			11.07.2023 10:19
M906/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	Chol2	244 mg/dL			11.07.2023 11:02
M910/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	Cl-C	95 mmol/L			11.07.2023 10:17
M906/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	Crea2	2.29 mg/dL			11.07.2023 11:22
M910/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	Fe-Pl	255 ug/dL			11.07.2023 10:28
M906/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	GGT	66 U/L			11.07.2023 11:57
M910/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	GluC	82.03 mg/dL			11.07.2023 10:20
M910/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	K-C	3.8 mmol/L			11.07.2023 10:17
M910/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	Na-C	148 mmol/L			11.07.2023 10:17
M910/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	Phos	3.67 mmol/L			11.07.2023 10:23
M910/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	TPRO2	6.4 g/dL			11.07.2023 10:22
M910/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	Trig2	196 mg/dL			11.07.2023 10:27
M910/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	UHDL	92.20 mg/dL			11.07.2023 10:26
M910/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	UIBC-L	78 ug/dL			11.07.2023 10:28
M916/1	5	BIORAD CHLevel 1	BIORAD CH - Level 1	UREA2	36.3800			11.07.2023 12:03
M910/1	1	BIORAD CHLevel 1	BIORAD CH - Level 1	Uric2	4.5 mg/dL			11.07.2023 10:26

Printed on: 11.07.2023
12:10:45PM

ARCHITECT

Page: 1 of 1

QC Results List Report

Operator ID: ADMIN

Released

System serial number: 86357

C / P	Module	SID	Control name	Assay	Result	Flags	Code	Date / Time
M910/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	AlbBCG2	2.6 g/dL			11.07.2023 10:23
M910/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	AlkP	410 U/L			11.07.2023 10:26
M910/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	ALT	95 U/L			11.07.2023 10:30
M910/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	AST	187 U/L			11.07.2023 10:29
M910/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	BiID	1.65 mg/dL			11.07.2023 10:29
M910/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	BiliT	4.39 mg/dL			11.07.2023 10:25
M910/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	CaC	13.31 mg/dL			11.07.2023 10:22
M906/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	Chol2	109 mg/dL			11.07.2023 11:02
M910/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	Cl-C	86 mmol/L			11.07.2023 10:21
M906/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	Crea2	5.72 mg/dL			11.07.2023 11:23
M910/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	Fe-PI	73 ug/dL			11.07.2023 10:31
M906/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	GGT	176 U/L			11.07.2023 11:57
M910/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	GluC	294.49 mg/dL			11.07.2023 10:23
M910/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	K-C	6.2 mmol/L			11.07.2023 10:21
M910/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	Na-C	124 mmol/L			11.07.2023 10:21
M910/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	Phos	7.30 mmol/L			11.07.2023 10:27
M910/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	TPRO2	4.2 g/dL			11.07.2023 10:25
M910/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	Trig2	88 mg/dL			11.07.2023 10:30
M910/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	UHDL	29.56 mg/dL			11.07.2023 10:29
M910/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	UIBC-L	118 ug/dL			11.07.2023 10:31
M906/2	5	BIORAD CHLevel 2	BIORAD CH - Level 2	UREA2	102.7200			11.07.2023 11:22
M910/2	1	BIORAD CHLevel 2	BIORAD CH - Level 2	Uric2	9.4 mg/dL			11.07.2023 10:30

Printed on: 11.07.2023
12:11:35PM

ARCHITECT

Page: 1 of 1