

HORIBA

Explore the future

HORIBA India Private Ltd.

246, Okhla Industrial

Estate, Phase - III 110

020 New Delhi, India

Tel: +91 (11) 4646 5000 / 4669 5001

Fax: +91 (11) 4669 5010 / 4646 5020

HIN/MED/2022-2023/100717

02nd Dec 2022

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer **ABX YUMIZEN H550** bearing serial number: **110YAXH03529** installed at **Tata 1MG Pvt Ltd. Dehradun** calibrated on **02nd Dec 2022**

Calibrator : ABX MINOCAL

Lot No. : CX477

Expiry Date : 05th Jan 2023

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on **01st Dec 2023**.



Shrish Dixit

(Head- Products & Marketing)

For Horiba India Pvt. Ltd.

Repeatability Report (part 1)

Number of repeatability run report selected for statistic calculation 6/16

Coefficient	WBC (10 ³ /μL)	RBC (10 ⁶ /μL)	HGB (g/dL)	HCT (%)	PLT (10 ³ /μL)	MCV (μm ³)	RDW-CV (%)	RDW-SD (μm ³)	P-LCR (%)
Minimum	7.89	4.94	9.7	31.0	278	62.6	15.9	35.3	35.4
Maximum	8.37	5.05	9.9	31.7	305	63.0	16.3	36.1	39.1
Mean	8.19	4.99	9.8	31.3	294	62.8	16.0	35.4	37.0
Difference	0.48	0.11	0.2	0.7	27	0.3	0.4	0.8	3.8
2 SD	0.38	0.08	0.2	0.6	18	0.2	0.3	0.7	2.6
CV(%)	2.34	0.84	1.07	0.92	3.02	0.19	0.85	0.97	3.46

Sei	Run Date & Time	WBC (10 ³ /μL)	RBC (10 ⁶ /μL)	HGB (g/dL)	HCT (%)	PLT (10 ³ /μL)	MCV (μm ³)	RDW-CV (%)	RDW-SD (μm ³)	P-LCR (%)	Operator
✓	12/02/2022 06:28:12 PM	8.28	4.97	9.7	31.1	296	62.6	15.9	35.3	36.5	1MG
✓	12/02/2022 06:33:22 PM	7.89	5.02	9.8	31.6	278	63.0	15.9	35.3	36.4	1MG
✓	12/02/2022 06:37:59 PM	8.28	4.94	9.7	31.0	297	62.9	16.3	36.1	37.6	1MG
✓	12/02/2022 06:48:37 PM	8.00	4.98	9.7	31.3	294	62.9	16.0	35.3	35.4	1MG
✓	12/02/2022 06:52:15 PM	8.37	4.96	9.9	31.1	305	62.7	16.0	35.3	37.3	1MG
✓	12/02/2022 06:54:42 PM	8.29	5.05	9.9	31.7	296	62.8	16.0	35.3	39.1	1MG
	03/30/2023 11:21:41 AM	6.84	4.98	13.1	40.2	361 *	80.6	13.3	37.8	28.3	1MG
	03/30/2023 11:23:14 AM	6.83	4.88	13.1	39.2	344	80.2	13.3	37.8	28.1	1MG
	03/30/2023 11:24:48 AM	6.97	4.87	13.4	39.0	341	80.1	13.3	37.8	30.1	1MG
	03/30/2023 11:26:25 AM	6.92	4.85	13.3	39.5	346	81.3	13.9	39.5	28.4	1MG
	03/30/2023 11:28:04 AM	6.87	4.89	13.1	39.6	348	81.0	13.6	38.6	28.7	1MG
	04/05/2023 10:25:09 AM	6.07	4.96	14.4	43.4	263	87.5	15.3	47.9	30.6	1MG
	04/05/2023 10:26:44 AM	6.28	4.94	14.4	43.0	254	87.0	15.1	47.0	28.7	1MG
	04/05/2023 10:28:19 AM	6.11	4.93	14.3	43.0	249	87.2	14.6	46.2	29.3	1MG
	04/05/2023 10:29:52 AM	6.04	4.94	14.4	43.1	247	87.2	14.7	46.2	31.3	1MG
	04/05/2023 10:31:40 AM	6.14	4.88	14.2	42.6	253	87.3	14.4	45.4	30.6	1MG

Repeatability Report (part 2)

Number of repeatability run report selected for statistic calculation 6/16

Coefficient	NEU% (%)	LYM% (%)	MON% (%)	EOS% (%)	BAS% (%)	LIC% (%)
Minimum	75.0	11.6	9.7	0.1	0.2	0.6
Maximum	77.7	12.8	12.3	0.2	0.5	0.9
Mean	76.5	12.3	10.7	0.2	0.4	0.7
Difference	2.8	1.2	2.6	0.1	0.3	0.3
2 SD	1.9	0.8	1.8	0.1	0.2	0.2
CV(%)	1.27	3.41	8.30	36.51	28.17	14.08

Selected	Run Date & Time	NEU% (%)	LYM% (%)	MON% (%)	EOS% (%)	BAS% (%)	LIC% (%)	Operator
✓	12/02/2022 06:28:12 PM	77.7	11.6	10.3	0.1	0.3	0.6	1MG
✓	12/02/2022 06:33:22 PM	76.8	12.1	10.5	0.1	0.5	0.7	1MG
✓	12/02/2022 06:37:59 PM	76.7	12.4	10.4	0.1	0.4	0.7	1MG
✓	12/02/2022 06:48:37 PM	75.8	12.6	11.0	0.2	0.4	0.9	1MG
✓	12/02/2022 06:52:15 PM	76.9	12.8	9.7	0.2	0.4	0.8	1MG
✓	12/02/2022 06:54:42 PM	75.0	12.3	12.3	0.2	0.2	0.7	1MG
	03/30/2023 11:21:41 AM	37.7	51.3	8.2	1.5	1.3	0.4	1MG
	03/30/2023 11:23:14 AM	38.3	50.1	8.5	1.7	1.4	0.2	1MG
	03/30/2023 11:24:48 AM	38.0	50.8	8.3	1.6	1.3	0.4	1MG
	03/30/2023 11:26:25 AM	38.6	49.5	8.3	1.8	1.8	0.3	1MG
	03/30/2023 11:28:04 AM	38.1	52.2	7.0	1.1	1.6	0.4	1MG
	04/05/2023 10:25:09 AM	50.6	39.1	6.6	2.3	1.4	0.2	1MG
	04/05/2023 10:26:44 AM	52.2	38.2	5.9	2.3	1.4	0.3	1MG
	04/05/2023 10:28:19 AM	50.9	38.6	6.3	2.7	1.5	0.3	1MG
	04/05/2023 10:29:52 AM	51.8	38.3	5.9	2.2	1.8	0.6	1MG
	04/05/2023 10:31:40 AM	51.1	39.6	6.2	2.0	1.1	0.6	1MG

Calibration Report

Sample ID CX477
Lot number CX477

Name ABX MINOCAL

Exp. date 01/05/2023
Modified on

Coefficient	WBC	RBC	HGB	HCT	PLT	MPV
New	1.136	0.996	0.988	1.042	1.078	1.028
Current	1.136	0.996	0.988	1.042	1.078	1.028
Target	8.78	4.62	13.6	39.3	270	11.1
Mean	8.82	4.64	13.6	39.2	263	11.3
CV(%)	1.19	0.68	0.82	0.90	3.54	0.96

Number of calibration run selected for coefficient calculation (minimum 5) 5/7

Sel.	Run Time	WBC (10 ³ /μL)	RBC (10 ⁶ /μL)	HGB (g/dL)	HCT (%)	PLT (10 ³ /μL)	MPV (μm ³)
✓	12/02/2022 12:57:52 PM	8.86	4.67	13.7	39.4	266	11.4
✓	12/02/2022 12:59:27 PM	8.72	4.67	13.5	39.7	275	11.1
✓	12/02/2022 01:01:06 PM	8.72	4.64	13.7	39.1	264	11.3
✓	12/02/2022 01:02:45 PM	8.97	4.63	13.7	39.2	250	11.3
	12/02/2022 01:04:24 PM	8.88	4.59	13.3	38.6	262	11.2
	12/02/2022 01:06:04 PM	8.81	4.55	13.1	38.5	258	11.4
✓	12/02/2022 01:08:08 PM	8.81	4.59	13.5	38.8	259	11.3

QC - Control Run Report

Run Date 12/02/2022 01:41:03 PM

Operator technician

Name PX438H


Sample ID PX438H

Level High


Exp. date 01/05/2023

Lot number PX438H


			Range
RBC	5.21	10 ⁶ /μL	5.04 - 5.54
HGB	15.7	g/dL	15.5 - 16.7
HCT	48.6	%	46.2 - 51.2
MCV	93.2	μm ³	87.0 - 97.0
MCH	30.2	pg	27.9 - 32.9
MCHC	32.4	g/dL	30.1 - 36.1
RDW-CV	13.7	%	10.5 - 18.5



			Range
PLT	493	10 ³ /μL	440 - 540
MPV	9.1	μm ³	7.0 - 11.0



			Range		Range
WBC	18.16	10 ³ /μL	15.70 - 20.10		
	#		Range	%	Range
NEU	12.81		10.50 -	70.5	59.3 - 79.3
LYM	3.87		2.29 - 5.29	21.3	13.2 - 29.2
MON	0.46		0.01 - 1.07	2.6	0.0 - 6.0
EOS	0.77		0.03 - 1.33	4.2	0.1 - 7.5
BAS	0.25		0.00 - 0.96	1.4	0.0 - 5.4



Alarms

Raw Data

PLT counts (by sections)

S1	S2	S3	S4	S5	S6
400	428	412	430	414	470
S7	S8	S9	S10	S11	S12
375	408	412	414	439	438

RBC counts (by sections)

S1	S2	S3	S4	S5	S6
4409	4336	4450	4495	4369	4372
S7	S8	S9	S10	S11	S12
4368	4379	4419	4437	4421	4387

WBC counts (by sections) and others

S1	S2	S3	S4	S5	S6
1284	1122	1168	1133	1140	1186
S7	S8	S9	S10	S11	
1171	1153	1204	1168	1178	

HBG measures (Optical Intensity)

LI1	LI2	LI3	LI4	LI5	LI6
3844	3845	3845	691	692	691
LI7	LI8	LI9	LI10	LI11	LI12
691	692	692	693	693	694
LI13					
693					

RMeasured : 10677

Correlated : 10519

Optical intensity 193

QC - Control Run Report

Run Date 12/02/2022 01:46:57 PM

Operator technician

Name PX438L


Sample ID PX438L

Level Low

Exp. date 01/05/2023


Lot number PX438

			Range
RBC	2.35	10 ⁶ /μL	2.18 - 2.50
HGB	6.2	g/dL	5.8 - 6.6
HCT	18.9	%	17.2 - 20.2
MCV	80.7	μm ³	75.0 - 85.0
MCH	26.3	pg	24.5 - 28.5
MCHC	32.5	g/dL	30.1 - 36.1
RDW-CV	17.2	%	13.0 - 21.0




Alarms

			Range
PLT	66	10 ³ /μL	45 - 85
MPV	9.4	μm ³	7.5 - 11.5



			Range		Range
WBC	3.30	10 ³ /μL	2.55 - 3.35		
	#		Range	%	Range
NEU	1.58		1.02 - 1.72	48.0	36.3 - 56.3
LYM	1.18		0.77 - 1.43	35.8	25.2 - 49.2
MON	0.20		0.01 - 0.43	6.0	0.0 - 15.0
EOS	0.30		0.00 - 0.38	9.0	0.3 - 12.7
BAS	0.04		0.01 - 0.13	1.2	0.4 - 4.6



Raw Data

PLT counts (by sections)

S1	S2	S3	S4	S5	S6
55	79	79	100	89	104
S7	S8	S9	S10	S11	S12
71	72	92	96	86	92

RBC counts (by sections)

S1	S2	S3	S4	S5	S6
2190	2197	2201	2213	2165	2153
S7	S8	S9	S10	S11	S12
2196	2148	2157	2120	2161	2126

WBC counts (by sections) and others

S1	S2	S3	S4	S5	S6
209	208	234	224	220	204
S7	S8	S9	S10	S11	S12
216	205	246	189	210	

RMeasured : 2141
Correlated : 2091
Optical intensity 193

HBG measures (Optical Intensity)

LI1	LI2	LI3	LI4	LI5	LI6
3846	3845	3846	1906	1905	1905
LI7	LI8	LI9	LI10	LI11	LI12
1905	1905	1905	1904	1905	1905
LI13					
1904					

QC - Control Run Report

Run Date 12/02/2022 01:48:45 PM

Operator technician

Name PX438L


Sample ID PX438L

Level Low

Exp. date 01/05/2023


Lot number PX438

			Range
RBC	2.30	10 ⁶ /μL	2.18 - 2.50
HGB	6.1	g/dL	5.8 - 6.6
HCT	18.5	%	17.2 - 20.2
MCV	80.4	μm ³	75.0 - 85.0
MCH	26.4	pg	24.5 - 28.5
MCHC	32.8	g/dL	30.1 - 36.1
RDW-CV	16.5	%	13.0 - 21.0




Alarms

			Range
PLT	70	10 ³ /μL	45 - 85
MPV	9.4	μm ³	7.5 - 11.5



			Range
WBC	3.09	10 ³ /μL	2.55 - 3.35

	#	Range	%	Range
NEU	1.53	1.02 - 1.72	49.5	36.3 - 56.3
LYM	1.09	0.77 - 1.43	35.3	25.2 - 49.2
MON	0.19	0.01 - 0.43	6.2	0.0 - 15.0
EOS	0.25	0.00 - 0.38	8.1	0.3 - 12.7
BAS	0.03	0.01 - 0.13	0.9	0.4 - 4.6



Raw Data

PLT counts (by sections)

S1	S2	S3	S4	S5	S6
82	103	90	87	88	93
S7	S8	S9	S10	S11	S12
75	70	79	93	92	101

RBC counts (by sections)

S1	S2	S3	S4	S5	S6
2097	2146	2210	2121	2120	2164
S7	S8	S9	S10	S11	S12
2191	2118	2071	2113	2087	2173

WBC counts (by sections) and others

S1	S2	S3	S4	S5	S6
192	206	208	229	171	208
S7	S8	S9	S10	S11	S12
186	199	193	209	216	

HBG measures (Optical Intensity)

L11	L12	L13	L14	L15	L16
3826	3845	3848	1924	1924	1924
L17	L18	L19	L110	L111	L112
1923	1923	1923	1923	1922	1922
L113					
1921					

RMeasured : 2114
Correlated : 2055
Optical Intensity : 193

Blank Cycle Logs

Running Date Comments	Operator	WBC 10 ³ /μL	RBC 10 ⁶ /μL	HGB g/dL	PLT 10 ³ /μL	Status	Technical alarms
12/02/2022 09:42:24 AM	1MG	0.00	0.00	0.0	0	Passed	
12/02/2022 09:50:33 AM	1MG	0.00	0.00	0.0	0	Passed	
12/02/2022 12:16:20 PM	1MG	0.00	0.00	0.0	1	Passed	

Notification No.	
Work Order No.	

SERVICE REPORT

CUSTOMER COPY

HIN-M02 **8337**

CUSTOMER DETAILS		INSTRUMENT DETAILS			SERVICE STATUS	
NAME:		MODEL : <i>Unitzen H350</i>			VISITING PURPOSE:	
ADDRESS:		EQUIP. SL. No. : <i>H09AXH03329</i>			<input type="checkbox"/> REPAIR <input type="checkbox"/> DATA ERROR <input type="checkbox"/> INSTALLATION <input type="checkbox"/> MAINTENANCE <input type="checkbox"/> UPGRADE <input type="checkbox"/> CUSTOMER TRAINING <input type="checkbox"/> COURTESY VISIT <input type="checkbox"/> DEMO <input type="checkbox"/> OTHERS	
CITY:		VERSION : <i>1.2.1-4</i>			CUSTOMER STATUS	
PIN CODE:		NOTIFICATION No.			<input checked="" type="checkbox"/> R.R <input type="checkbox"/> WARRANTY <input type="checkbox"/> AMC <input type="checkbox"/> CMC <input type="checkbox"/> DEMO <input type="checkbox"/> FREE SERVICE <input type="checkbox"/> CHARGEABLE CALL <input type="checkbox"/> OTHERS	
CITY:		CALL DETAILS			WORKS CARRIED OUT AT	
DATE:		DATE			<input type="checkbox"/> SITE <input type="checkbox"/> SERVICE CENTRE	
TEL. NO.:		TIME				
CONTACT PERSON :		COMPLAINT RECD.				
PROBLEM REPORTED :		START				
OBSERVATIONS :		COMPLETED				
ACTION TAKEN :		TRAVEL TIME (Eng. 1)			HOURS	
		TRAVEL TIME (Eng. 2)			HOURS	
		DAILY WORKLOAD				
<input type="checkbox"/> FOLLOWING PARTS HAVE BEEN REPLACED <input type="checkbox"/> FOLLOWING PARTS NEED TO BE REPLACED PLEASE APPROVE						
PART CODE.	DESCRIPTION (Replaced)	QTY.	COST	TAX	TOTAL	
	<i>CX 477</i>					
	<i>Expiry - 05/01/2023</i>					
PART CODE.	DESCRIPTION (Trouble Shooting)	QTY.	TOTAL RS.			
				*Replaced for Trouble Shooting Need to be returned after Trouble Shooting.		
FOLLOW-UP ACTION (Spare need to replaced, if any)			TO BE FILLED IN BY CUSTOMER			
			<input type="checkbox"/> FAULT RECTIFIED & INSTRUMENT IS WORKING SATISFACTORILY <input type="checkbox"/> FAULT IS NOT COMPLETED, ENGINEER NEED TO COME AGAIN <input type="checkbox"/> WE HEREBY APPROVE Rs. <input type="checkbox"/> COMMENTS (IF ANY)			
ENGINEER'S NAME-1:		ENGINEER'S NAME-2:		CUSTOMER SIGNATURE:		
SIGNATURE:		SIGNATURE:		NAME :		
DATE & TIME:		DATE & TIME:		SEAL:		