



# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

|                             |  |                 |                     |                       |
|-----------------------------|--|-----------------|---------------------|-----------------------|
| CERTIFICATE NO.             | : 221005A/N-116  | Page            | : 1 of 1            |                       |
| CALIBRATED ON.              | : 06 / 10 / 2022   | Ref. Std.       | : CG-11             |                       |
| CALIBRATION DUE.            | : 05 / 10 / 2023   | C.P. No.        | : ACS/CP/07         |                       |
| SRF No. & DATE              | : 221005A/N Dt. 05.10.2022   | Phy. Inspection | : O.K.              |                       |
| ULR No. & ISSUE DATE        | : CC342722000000629F Dt. 14.10.2022  | Cal. At         | : Site / Lab        |                       |
| CALIBRATED FOR              | : M/s Redcliffe Lifetech Pvt. Ltd.<br>Plot No. 27A/37/38, Near HDFC Bank,<br>Shastri Circle, Jodhpur, 342003 |                 |                     |                       |
| ITEM CALIBRATED             | : Digital Temp. Indicator  |                 |                     |                       |
|                             | Make : RoHS  | Range           | : (-50 to 110) °C   |                       |
|                             | L. Count   | ID No.          | : RCL/HEM/D.TEMP-01 |                       |
|                             | Location : Hematology  | S. No.          | : N.A               |                       |
| ENVIRONMENT CONDITION:      | Temp. { 25°C ± 3°C} R.H. {40% to 60%}  |                 |                     |                       |
| STANDARD INSTRUMENTS USED : |  |                 |                     |                       |
| INSTT. NAME                 | ID No. / S. No.  | TRACEABILITY    | CERTIFICATE No.     | DUE DATE              |
| Digital Calibrator          | ACS/UC/01 / 300561   | Anshaankan      | AIPL/2021/0784      | 11/11/2022            |
| RESULTS                     | : ET Discipline  |                 |                     |                       |
| * U. C. INSTT. READS (°C)   | STANDARD READING (°C)  | ERROR (°C)      |                     |                       |
| 2.02                        | 2  | 0.02            |                     |                       |
| 4.04                        | 4  | 0.04            |                     |                       |
| 6.04                        | 6  | 0.04            |                     |                       |
| 8.06                        | 8  | 0.06            |                     |                       |
|                             |  |                 |                     | (average of readings) |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=2.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

