



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name **DISTRICT HOSPITAL, JABALPUR** Lab No **7821**
Constituent Group **HbA1c** Date of Result Entered : **17/06/2023**
PT item **Lyophilized human whole blood based** Date of Report Published : **04/07/2023**

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|---------|-------------------------|---------------------------------------|--------------------|------|--------------|------|------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | HbA1c | HPLC-Ion Exchange | Biorad (D10 / D100 / Variant Turbo) | 715 | 6.80 | 2.65 | 0.18 | 6.7 % | -0.56 | 0.01 |

| SDI Range | Interpretation |
|---------------------------------|------------------------------------------|
| Within -1.00 to +1.00 | Excellent. |
| Within ± 1.01 to ± 2.00 | Good. |
| Within ± 2.01 to ± 2.99 | Accept with caution. Warning Signal. |
| Beyond ± 3.0 | Unacceptable performance. Action Signal. |

LAB ADDRESS :

DISTRICT HOSPITAL, JABALPUR
SETH GOVIND DAS(VICTORIA) HOSPITAL, JABALPUR
JABALPUR
MADHYA PRADESH 482008

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****