

## BIO-RAD LABORATORIES (INDIA) PRIVATE LIMITED

| MADE VICTORIA HOSDITAL |
|------------------------|
| ASH VICTORIA HOSPITAL  |
| AJABALPUR              |
|                        |
|                        |
| Ŀ                      |

## **Calibration Certificate**

| Date | 15.10.2022      | Instrument |                   |                 |            |
|------|-----------------|------------|-------------------|-----------------|------------|
| Date | 13.10.2022      | Ansu ument | D-10 Hb Applican  | Instrument      | DJOB475808 |
|      | (dd mmmm, yyyy) | Model      | B-10 110 Analyzer | Serial Number   | DJOB475808 |
| 147  |                 |            |                   | Scrial Mulliber |            |

We certify that the Bio-Rad instrument specified above and installed in your lab is calibrated and the performance was found satisfactory.

High Pressure Pump Flow RateReference: Service Manua

| Measured with (Device manufacturer and model Measuring cylinder hex base 10ML | Device Serial Number 2310100 |                | Device Calibration date  (dd mmmm, yyyy)  09.09.22 |  |
|---|------------------------------|----------------|--|--|
| Specified Range@ 1.5ml/M  | inute                        | Measured Value |  |  |
| 4 40 -  |                              | 1.55           |  |  |

Column Holder temperature Reference: Service Manual: Service M

| Measured with (Device manufacturer and model | Device Serial Number | Device Calibration date  (dd mmmm, yyyy) |  |
|--|----------------------|--|--|
| FLUKE 179                                    | 26030078             | 06.06.22                                 |  |
| Specified Range @ 35°C                       | Measured Value       |  |  |
| 33.4°C – 36.4°C                              | 34.4                 |  |  |

**Detector log value** 

Reference: Service Manual: Section 7.2 or Field Bulletin D-10-081

| Measured with               | Specified Range  | Measured Value |  |  |
|-----------------------------|------------------|----------------|--|--|
| Instrument Service Software | 0.018 - 0.022 AU | 0.021AU        |  |  |

14.10.2023 The next calibration is due on (dd mmmm, yyyy)

For Bio-Rad Laboratories India Pvt Limited

Authorized Signatory:

**Calibration Certificate** 

Page 1 of 1