



ARSH ENTERPRISE

(NABL Accredited Laboratory as per ISO/IEC 17025)

A House Of Calibration

A-337, Sumel Business Park-4, Kalupur-Naroda Road, Amdupura-382345, Ahmedabad.
M. 9974712437, 9173769174 | E-mail : arshenterprise1992@gmail.com



Calibration Certificate

Discipline : Thermal ULR/Certificate No.: CC312623000015597F

1) Customer Information:

Customer Details:	Seth Govinddas (Victoria) District Govt.	SRF No.	AE/SRF/23/1825
	Hospital	Receipt Date	28/06/2023
	Jabalpur Central Processing Lab	Calibration Date	28/06/2023
	Badi Omti Gangipura Road, Jabalpur-482002,	Due Date	27/06/2024
	MP.	Date of Issue	30/06/2023

2) Instrument Information:

Calibrated Item	Refrigerator	Visual Inspection	Ok
Make/Model	Trufrost / ===	Type	===
Range	0 - 10 °C	Least Count	1°C
Serial No./ID No.:	FKG400 / VHJ/CPL/REF/24	Calibration Performed	Site
User Location	Lab	Parameter	Temperature

3) Environmental Conditions:

Temperature	25 ± 2 °C	Calibration Ref Std No.	ITS-90
Relative Humidity	50 ± 20 % RH	Calibration Procedure No.	AE/CAL/T/SOP-03

4) Reference Standards Used:

Reference Standard	Serial No.	Certificate No.	Valid Up To	Traceability
4 Wire RTD PT-100	1216	2207/0715/01	01/07/2023	CC-2597
Multifunction Calibrator	994413890	AMC/00000226	13/02/2024	CC-2364

5) Calibration Results:


Set Reading On UUC	Reading Observed on Master	Error (±)	Expanded Uncertainty (±°C)
2.0 °C	2.0 °C	0.0 °C	1.07
4.0 °C	4.0 °C	0.0 °C	
6.0 °C	5.9 °C	-0.1 °C	
8.0 °C	7.8 °C	-0.2 °C	
10.0 °C	9.8 °C	-0.2 °C	

6) Notes:

1. For All Notes Kindly Refer Overleaf

Calibrated By

Yash Khatri
(Calibration Engineer)



Authorised Signatory

Mansuri Md.Arif
(Quality Manager)

- End of Certificate -