



ARSH ENTERPRISE

(NABL Accredited Laboratory as per ISO/IEC 17025)

A House Of Calibration

A-337, Sumel Business Park-4, Kalupur-Naroda Road, Amdupura-382345, Ahmedabad.
M. 9974712437, 9173769174 | E-mail : arshenterprise1992@gmail.com



Calibration Certificate

Discipline : Mechanical Certificate No.: CC312623000015595F

1) Customer Information:

Customer Details:	Seth Govinddas (Victoria) District Govt. Hospital	SRF No.	AE/SRF/23/1825
	Jabalpur Central Processing Lab	Receipt Date	28/06/2023
	Badi Omti Gangipura Road, Jabalpur-482002, MP.	Calibration Date	28/06/2023
		Due Date	27/06/2024
		Date of Issue	30/06/2023

2) Instrument Information:

Calibrated Item	Centrifuge Machine	Visual Inspection	OK
Make/Model	Laby / ===	Type	Digital
Range	5000 RPM	Least Count	1 RPM
Serial No./ID No.:	2019-20/137 / VHJ/CPL/CENTRIFUGE/20	Calibration Performed	At Site
User Location	Lab	Parameter	Speed

3) Environmental Conditions:

Temperature	25 ± 5 °C	Calibration Ref Std No.	===
Relative Humidity	50 ± 20 % RH	Calibration Procedure No.	AE/CAL/M/SOP-27

4) Reference Standards Used:

Reference Standard	Serial No.	Certificate No.	Valid Up To	Traceability
Digital Tachometer	AE/M/28	PE23CC02645	22/05/2024	CC-3034

5) Calibration Results:


Reading Observed in UUC	Reading shown in Master Digital Tachometer	Error(±)
3000 RPM	2995.0 RPM	-5.0 RPM
3500 RPM	3493.7 RPM	-6.3 RPM
4000 RPM	3989.1 RPM	-10.9 RPM
5000 RPM	4987.9 RPM	-12.1 RPM

6) Notes:

1. For All Notes Kindly Refer Overleaf

Calibrated By

Yash Khatri
(Calibration Engineer)



Authorised Signatory

Mansuri Md. Arif
(Quality Manager)

- End of Certificate -