



# ARSH ENTERPRISE

(NABL Accredited Laboratory as per ISO/IEC 17025)

## A House Of Calibration

A-337, Sumel Business Park-4, Kalupur-Naroda Road, Amdupura-382345, Ahmedabad.  
M. 9974712437, 9173769174 | E-mail : arshenterprise1992@gmail.com



### Calibration Certificate

<b>Discipline :</b>	Mechanical	<b>Certificate No.:</b>	CC312623000015596F	
<b>1) Customer Information:</b>				
Customer Details:	<b>Seth Govinddas (Victoria) District Govt. Hospital</b>	SRF No.	AE/SRF/23/1825	
	Jabalpur Central Processing Lab	Receipt Date	28/06/2023	
	Badi Omti Gangipura Road, Jabalpur-482002, MP.	Calibration Date	28/06/2023	
		Due Date	27/06/2024	
		Date of Issue	30/06/2023	
<b>2) Instrument Information:</b>				
Calibrated Item	<b>Centrifuge Machine</b>	Visual Inspection	Ok	
Make/Model	Laby / ===	Type	Digital	
Range	5000 RPM	Least Count	1 RPM	
Serial No./ID No.:	=== / VHJ/CPL/CENTRIFUGE/30	Calibration Performed	At Site	
User Location	Lab	Parameter	Speed	
<b>3) Environmental Conditions:</b>				
Temperature	25 ± 5 °C	Calibration Ref Std No.	===	
Relative Humidity	50 ± 20 % RH	Calibration Procedure No.	AE/CAL/M/SOP-27	
<b>4) Reference Standards Used:</b>				
<b>Reference Standard</b>	<b>Serial No.</b>	<b>Certificate No.</b>	<b>Valid Up To</b>	<b>Traceability</b>
Digital Tachometer	AE/M/28	PE23CC02645	22/05/2024	CC-3034
<b>5) Calibration Results:</b>				
Reading Observed in UUC	Reading shown in Master Digital Tachometer	Error(±)		
3000 RPM	2998.1 RPM	-1.9 RPM		
3500 RPM	3497.9 RPM	-2.1 RPM		
4000 RPM	3994.9 RPM	-5.1 RPM		
5000 RPM	4991.7 RPM	-8.3 RPM		
<b>6) Notes:</b>				
1. For All Notes Kindly Refer Overleaf				
Calibrated By			Authorised Signatory	
 Yash Khatri (Calibration Engineer)			 Mansuri Md.Arif (Quality Manager)	
- End of Certificate -				