



# ARSH ENTERPRISE

(NABL Accredited Laboratory as per ISO/IEC 17025)

## A House Of Calibration

A-337, Sumel Business Park-4, Kalupur-Naroda Road, Amdupura-382345, Ahmedabad.  
M. 9974712437, 9173769174 | E-mail : arshenterprise1992@gmail.com



### Calibration Certificate

Discipline : Mechanical ULR/Certificate No.: CC312623000015593F

#### 1) Customer Information:

Customer Details:	<b>Seth Govinddas (Victoria) District Govt. Hospital</b>	SRF No.	AE/SRF/23/1825
	Jabalpur Central Processing Lab	Receipt Date	28/06/2023
	Badi Omti Gangipura Road, Jabalpur-482002, MP.	Calibration Date	28/06/2023
		Due Date	27/06/2024
		Date of Issue	30/06/2023

#### 2) Instrument Information:

Calibrated Item	<b>Micro Pipette</b>	Visual Inspection	Ok
Make/Model	Dragon / ===	Type	===
Range	5 to 50 µl	Least Count	0.1 µl
Serial No/ ID NO:	YE202AM0110352 / VHJ/CPL/DRAGON/19	Calibration Performed	At Lab
User Location	Lab	Parameter	Volume

#### 3) Environmental Conditions:

Temperature	20 ± 2 °C	Calibration Ref Std No.	ISO 8655
Relative Humidity	50 ± 10 % RH	Calibration Procedure No.	AE/CAL/M/SOP-18

#### 4) Reference Standards Used:

Reference Standard	Serial No.	Certificate No.	Valid Up To	Traceability
Digital Analytical Weighing Balance	20181221	EWS-1022-082	18/10/2023	CC-3295

#### 5) Calibration Results:


Nominal Value On UUC	Reading Observed By Master	(±) Error
20 µl	19.19 µl	-0.81 µl
30 µl	28.81 µl	-1.19 µl
50 µl	48.54 µl	-1.46 µl

#### 6) Notes:

- For All Notes Kindly Refer Overleaf
- Expanded Uncertainty : ± 150 µl

Calibrated By

Anas Mansuri  
(Calibration Engineer)



End of Certificate -

Authorised Signatory

Mansuri Md.Arif  
(Quality Manager)