



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 1573

Distribution No.: 160-C

Month/Year: May/2023

Instrument ID: MINDRAY BC2800

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 11-07-2023[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	10.5	10.5	21	17.4	0.112	1.14	0	0.18	0.010	-1.07
RBC x10 <sup>6</sup> /µl	1	4.04	3.98	8.02	8.58	0.008	-2.36	0.06	0.04	0.002	0.54
Hb g/dl	1	11.7	11.6	23.3	25.1	0.026	-2.43	0.1	0.1	0.007	0.00
HCT%	1	37.1	36.3	73.4	75.3	0.191	-0.32	0.8	0.3	0.022	1.35
MCV-fl	1	91.9	91.4	183.3	175.3	0.389	0.68	0.5	0.3	0.018	0.67
MCH-Pg	1	29.1	28.9	58	58.6	0.077	-0.32	0.2	0.2	0.014	0.00
MCHC-g/dl	1	31.9	31.5	63.4	66.5	0.168	-0.61	0.4	0.3	0.016	0.34
Plt. x10 <sup>3</sup> /µl	1	118	94	212	200	2.329	0.17	24	6	0.342	3.04
Retic %	2	1	1	2	18.6	0.269	-2.13	0	0.5	0.031	-0.84

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT	
DLC%	3	Nrbcs= , Poly= L=30, E=, Mono/Promono= , B1=20 P.M.=, Mye=, Meta=, Other=50	Blast: 60-87, Lympho: 9-23, Poly: 1-4, nRBC/ Mono/Eos/Baso/Myelo/Meta/ Promyelo: 0-5	
RBC Morphology	3	MICROCYTIC HYPOCHROMICWITH ANISOPOIKILOCYTOSIS	Predominantly: Normocytic/ Normochromic, Moderate: Anisocytosis, Microcytic	
Diagnosis	3	CHRONIC LYMPHOCYTIC LEUKEMIA	Acute Leukemia (AL)	

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 160--C	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	366	362	80.11	87.02	9.12	4.97	10.77	8.01
<b>RBC x10<sup>6</sup>/µl</b>	1	366	366	88.52	92.08	5.46	4.37	6.02	3.55
<b>Hb g/dl</b>	1	366	366	84.43	89.34	7.65	5.74	7.92	4.92
<b>HCT%</b>	1	366	363	94.21	95.04	4.13	1.65	1.66	3.31
<b>MCV-fl</b>	1	366	363	93.11	88.43	4.68	6.89	2.21	4.68
<b>MCH-Pg</b>	1	366	363	81.54	90.91	7.44	4.41	11.02	4.68
<b>MCHC-g/dl</b>	1	366	363	93.94	86.5	3.58	3.86	2.48	9.64
<b>Plt. x10<sup>3</sup>/µl</b>	1	366	363	94.21	87.88	4.13	6.89	1.66	5.23
<b>ReticCount%</b>	2	366	342	93.27	86.84	4.09	10.23	2.64	2.93
<b>PS Assessment</b>	3	366	339	Satisfactory :83.62%, Borderline Sat. :9.83%, Unsatisfactory :6.55%					

**\*Comments:**

1). **Among Lab (EQA) : PS Diagnosis wrongly reported, remaining results acceptable**

2). **Within Lab (IQA) : Difference in the CBC measurement values for PLT unacceptable, may be due to random/human error.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----