



# Metropolis Healthcare Ltd., EQAS

RAHA DIAGNOSTICS (824)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 220412 and Sample No 02

Report Date : 25/01/2023

Analyte	Instrument	Result Value	Standard Unit
# HCV	Manual	Reactive	-
X Dengue IgG	Manual	Positive	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ Delayed Result Entry			

Problem Classification: \_\_\_\_\_

\_\_\_\_\_  
\_\_\_\_\_

Corrective Action: \_\_\_\_\_

\_\_\_\_\_  
\_\_\_\_\_

Reviewed by: \_\_\_\_\_

Dated: \_\_\_\_\_

Instrument : Manual


Analyte	Standard Unit	Result Value	Accepted Value
@ HBsAg	IU/mL	Reactive	Reactive
# HCV	IU/mL	Reactive	
@ RPR (Syphilis Serology)	-	Non-reactive	Non-reactive
@ Dengue NS 1	RU/ml	Negative	Negative
@ HIV Ab	S/CO	HIV-1 Ab Reactive	HIV-1 Ab Reactive
@ Dengue IgM	RU/ml	Negative	Negative
X Dengue IgG	RU/ml	Positive	Negative

Instrument : Erba EM 200

Analyte	Standard Unit	Result Value	Accepted Value
@ C-Reactive Protein Qualitative	mg/L	Abnormal	Abnormal
@ Rheumatoid Factor Qualitative	IU/mL	Normal	Normal
@ ASO Qualitative	IU/mL	Normal	Normal

Legend @ : Acceptable  
 @ : Acceptable  
 X : Unacceptable  
 # : Not Evaluated  
 ⊙ : Delayed Result Entry

Total Parameters	10
Not Evaluated Parameters	1
Evaluated Parameters	9
Outlier Parameters (X)	1
EQAS Score Immunoassay Extended	88.89 %



**Dr Puneet Kumar Nigam**  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kirol Road, Kurla (W),  
 Mumbai - 400070