



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name **Govt. Block PHC, Thimiri** Lab No **3562**  
 Constituent Group **Chemistry III** Date of Result Entered : **19/04/2023**  
 PT item **Lyophilized human serum based** Date of Report Published : **06/05/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	<b>103 mg/dL</b>	-2.28	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	<b>2 mg/dL</b>	-0.91	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	<b>120 mg/dL</b>	1.45	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

**Govt. Block PHC, Thimiri**  
**VELLORE DISTRICT**  
**Thimiri**  
**TAMILNADU632512**

**Coordinator Contact Details:**  
**Email:clinqc@cmcvellore.ac.in**  
**Contact Number: 0416-2283102**

*Pamela Christudoss*

**Dr. Pamela Christudoss**  
**CMC EQAS Coordinator**  
**Christian Medical College, Vellore**

**Homogeneity and Stability of the sample is passed.**  
**Data in CMC EQAS reports is confidential**