



KOMMUNJE HEALTHCARE



KOMMUNJE HEALTHCARE

Kommunje House, Manila Village, Post : Murva - 574243 Mob. : 7406031421 , 8088314654

CALIBRATION REPORT

Customer Name: BM laboratory and Digital X-ray
Equipment Name: Easy LYTE plus Electrolyte ANALYSER
Make: Medica Corporation
Serial Number: NA Asset ID :BML/EA/01
Date of Calibration: 24-03-2023 Calibration Due date: 24-03-2024

- 1.ISE Electrode condition solution -checked
- 2.Probe blockage -checked
- 3 Orings checked - Replaced
4. Peristaltic pump & Valves -Checked and Tubings Replaced
- 5.Liniarity & Precision Checked - Found Ok

Engineer Name: Poadeep Kumar Customer Name: Jithendra
Signature:  Sign & Stamp: 
Date : 24/03/23 Date : 24/3/23.

KOMMUNJE HEALTHCARE

M:8088314654

Manila Village Post Murva Bantwal
Taluk 574243





KOMMUNJE HEALTHCARE

KOMMUNJE HEALTHCARE

Kommunje House, Manila Village, Post : Murva - 574243 Mob. : 7406031421 , 8088314654

Linearity and SD and CV checked

Control: Biorad Lot Number :26490 Expiry :30-11-2023

Parameter: Sodium Mean value :143.3 Kit Insert SD :2.01 CV :1.4

1	143.2
2	142.8
3	142.9
4	143.5
5	143.5
6	142.9
7	142.8
8	143.2
9	143.1
10	143.1
11	142.9
12	143.2
13	142.9
14	142.9
15	143.2
16	143.2
17	143.1
18	143.4
19	143.5
20	143.1

Calculated Mean:143.1 SD :0.23 CV:0.16

Laniary Check Passed

