



## CALIBRATION CERTIFICATE

This is to certify that the cobas pure is a fully automated analyzer bearing the serial numbers **SSU: 2265-09, Cobas c 303:2256-10, e 402: 2254-09**, at **REDCLIFFE LIFE TECH PVT LTD, Hyderabad** has been calibrated on **11/07/2023**.

### The calibration includes:

**Adjustments:** Checked and adjusted sample/Reagent Rotor, Incubator, Sample/Reagent probe, Bead Mixer and their respective home positions.

Incubation Disk: Range: 36.8°C to 37.2°C  
Adjusted: 37°C

Detection Unit: Target: 28.0°C  
Adjusted: 28.0°C

PC/CC: Target: 28°C  
Adjusted: 28°C

Adding to this all the temperatures, Blank Cell Calibration and volumes drawn by all pumps were checked and found they are OK.

Next calibration due on 10/07/2024.

**For Roche Diagnostics India Pvt Ltd.,**

T R Siddardha,  
Sr. Technical Service Specialist,  
Hyderabad.



## Installation Qualification:

This document forms the basis of the Qualification Services Certificate. It certifies that the instrument is installed according to the manufacturer's specifications. The report presents and documents the test procedures, the documentation, reference and acceptance criteria used to verify that the system is installed according specifications. The report demonstrated that all installation qualification criteria have been met satisfactorily.

**Notice:** The following tests are to be carried out by trained Roche personnel only.

**Purpose:** The purpose of this test is to confirm that the instrument was delivered undamaged and installed correctly.

Test #	Test	Pass Fail	Signature Date
IQ.1.1	User assistance available	Pass	} Bmw 27/07/2022
IQ 1.2	Environmental parameters met	Pass	
IQ 1.3	Instrument delivered undamaged and complete	Pass	
IQ 1.4	Transport locking successfully removed	Pass	
IQ 1.5	All connections correctly installed	Pass	
IQ 1.6	Instrument positioned according to Installation Manual	Pass	
IQ 1.7	Instrument boot process successfully	Pass	
IQ 1.8	Checksum according to specification	Pass	
IQ 1.9	Mechanical adjustments complete	Pass	
IQ 1.10	Auxiliary components positioned	Pass	
IQ 1.11	Instrument installation check	Pass	
IQ 1.12	Host communication settings checked	Pass	

Test #	Test	Pass Fail	Signature Date
IQ.2	Installation Qualification for cobas <ISE>	Pass	} Bmw 27/07/2022
IQ.3	Installation Qualification for cobas <c 303>	Pass	
IQ.4	Installation Qualification for cobas <e 402>	Pass	
IQ.5	Installation Qualification for cobas link	Pass	



*Deviation Report:* Any discrepancies found during the installation must be documented in the space below. Roche personnel will then investigate the deviation and decide upon the most appropriate action to be taken.

Deviation #1

Investigation

Action taken

Deviation resolved satisfactorily? specify

N/A

Deviation #2

Investigation

Action taken

Deviation resolved satisfactorily? specify

N/A

Deviation #3

Investigation

Action taken

Deviation resolved satisfactorily? specify

N/A



## Operational Qualification:

This document is the basis of the Qualification Service Certificate. It certifies that the instrument is operating according to the manufacturer's specifications. This report presents and documents the test procedures, documentation, references and acceptance criteria used to verify that the specified system is operating according to the specifications. The report demonstrates that all operational qualification criteria have been met satisfactorily.

**Notice:** The following tests are to be carried out by trained Roche personnel only.

**Purpose:** The purpose of this test is to check that the modules are operating in accordance with the

Test #	Test	Pass Fail	Signature Date
OQ.1	Calibration successfully	Pass	} BMZ #10/9/2022
OQ.2	Quality Control successfully	Pass	
OQ.3	Accuracy check successfully	Pass	

**Deviation Report:** Any discrepancies found during the installation must be documented in the space below. Roche personnel will then investigate the deviation and decide upon the most appropriate action to be taken.

Deviation #1

Investigation

Action taken

Deviation resolved satisfactorily? specify

NA

Deviation #2

Investigation

Action taken

Deviation resolved satisfactorily? specify

NA





## cobas® pure integrated solutions

### Description

IQ.1.1	User assistance available	
	Check that a User Assistance opens and has content.	Pass
IQ.1.2	Environmental parameters	
	Ambient temperature in the lab is between 18° and 32 °C	Pass
	Ambient humidity at the lab is between 30 and 85% RH and non-condensing	Pass
	Bacteria free, deionized water < 10 cfu/ml	Pass
	Water conductivity 1.0 µS/cm or less	Pass
	Water pressure between 50 kPa and 340 kPa	Pass
	Instrument is not exposed to direct sunlight	Pass
	Floor is level and grade is ≤ 1/200 ( ≤0.5%)	Pass
IQ.1.3	Instrument delivered undamaged and complete	
	All covers are undamaged	Pass
	All accessory boxes are delivered	Pass
	Instrument does not show any external damage	Pass
IQ.1.4	Transport locking successfully removed	
	Unpacking of the different modules and accessories without damage to units	Pass
IQ.1.5	All connections correctly installed	
	Power distribution board and water supply/drainage facilities located within 5m from the instrument.	Pass
	Power supply voltage at the customer facility:	Pass
	UPS system available:	Pass
	Voltage fluctuation less than ±20V	Pass
	Grounding terminal of 10Ω or less available	Pass



IQ 1.6	Instrument positioned according to Installation Manual	
	System layout is according to the description in the manual	Pass
	Modules are installed according to the installation manual with official tools	Pass
IQ 1.7	Instrument boot process successful	
	IP address configuration correct	Pass
	First system boot-up	Pass
	Change cobas link IP Internet NIC (162.132.241.10)	Pass
IQ 1.8	Checksum according to specification	
	Version of installed cobas pure user software	
	Installation of country language successful	Pass
	Checksum of installed software is correct according to Installation Guide	Pass
IQ 1.9	Mechanical adjustments complete	
	All mechanical adjustments for the Sample Line and Rotor are carried out	Pass
	Rack transport during mechanical check function	Pass
	Mechanical adjustments backed up	Pass
IQ 1.10	Auxiliary components positioned	
	Rack trays are installed	Pass
IQ 1.11	Instrument installation check	
	Print function	Pass
	Download parameters from CL to CU PC	Pass
	Download applications	Pass
	Registered electrodes for ISE	Pass
	Rack/Sample barcode read check (attached printout)	Pass



IQ 1.12 Host communication settings checked

Host settings customised to local site and tested

Pass

**Deviation Report:** Any discrepancies found during the installation must be documented in the space below. Roche personnel will then investigate the deviation and decide upon the most appropriate action to be taken.

Deviation #1

Investigation

Action taken

Deviation resolved satisfactorily? specify

NA

Deviation #2

Investigation

Action taken

Deviation resolved satisfactorily? specify

NA

Deviation #3

Investigation

Action taken

Deviation resolved satisfactorily? specify

NA





## Installation Qualification for cobas® pure <ISE>

### Description

IQ.2.2	Mechanical adjustments complete	
	All mechanical adjustments for ISE mechanical parts are carried out	Pass
	Adjustment check during mechanical check function	Pass
IQ.2.3	Auxiliary components positioned	
	ISE Reagents are loaded	Pass
IQ.2.4	Gear pump adjustment	
	Gear pump adjustment executed (attached printout)	Pass
IQ.2.5	Instrument installation check	
	ISE Check 20 times (attached printout)	Pass
IQ.2.6	Application installation	
	Download of applications from cobas link: Na (29070) K (29080) Cl (29090)	Pass

**cobas®**



**Deviation Report:** Any discrepancies found during the installation must be documented in the space below. Roche personnel will then investigate the deviation and decide upon the most appropriate action to be taken.

Deviation #1

Investigation

Action taken

Deviation resolved satisfactorily? specify

NA

Deviation #2

Investigation

Action taken

Deviation resolved satisfactorily? specify

NA



## Installation Qualification for cobas® pure <c 303>:

### Description

IQ.3.1	Function check of c 303 module according to specifications	
	System layout is according to the description in the manual	Pass
	c 303 AU is installed according to the installation manual with official tools	Pass
IQ.3.2	Mechanical adjustments complete	
	All mechanical adjustments for the different c 303 mechanical parts are carried out	Pass
IQ.3.3	Probes and consumables installation	
	Sample Probe and Reagent probe are installed	Pass
	Reaction cells are installed	Pass
IQ.3.4	Instrument installation check	
	Basic and Acid wash bottles are loaded	Pass
	Water pressure: Main pump 50.0–60.0 kPa, Gear Pump 320 kpa	Pass
	Load ECO-D c pack green	Pass
	Exchange incubation bath water	Pass
	Water flow of rinse stations as well as consumption of the detergents were adjusted	Pass
	Air purge for syringes and reagents	Pass
	Photometer check (attached printout)	Pass
	Coil Blank Measurement (attached printout)	Pass
	Incubation water bath temperature 37 °C ± 0.1 °C	Pass



	Adjustment check during mechanical check function	Pass
IQ 3.5	Gear pump adjustment	
	Gear pump adjustment (attached printout)	Pass
IQ 3.6	Application installation	
	Download Special Wash (all)	Pass
	Download of Auxiliary Reagents from cobas link ECO-D, NAQHD, SMS and PYP (PYP if ASTP2 is not available) Menu>System>Auxiliary Reagent Packs>Download	Pass
	Download of applications from cobas link ASTP (20220) (if ASTP2 (20230) is not available), CHOL2 (20411) CREJ2 (20470) GLUC3 (20630) TP2 (21110) CONA-P2 (20993) CONA-R1 (21280) INST-S1 (21290) INST-R1 (21291) Menu>Application>Download	Pass
IQ 3.7	Instrument check	
	Instrument Check (attached printout)	Pass
IQ 3.8	Backup of adjustment data	
	Adjustments data backed up	Pass



**Deviation Report:** Any discrepancies found during the installation must be documented in the space below. Roche personnel will then investigate the deviation and decide upon the most appropriate action to be taken.

Deviation #1

Investigation

Action taken

Deviation resolved satisfactorily? specify

NA

Deviation #2

Investigation

Action taken

Deviation resolved satisfactorily? specify

NA



## Installation Qualification for cobas® pure <e 402>:

### Description

IQ.4.1	Function check of e 402 module according to specifications	
	e 402 AU is installed according to the installation manual with official tools	Pass
IQ.4.2	Mechanical adjustments complete	
	All mechanical adjustments for the different e 402 mechanical parts are carried out	Pass
	Adjustment check during mechanical check function	Pass
	Water pressure: Gear pump 320 kPa. Main pump 50.0–60.0 kPa	Pass
	Water flow of all rinse stations and wash station was adjusted and validated	Pass
IQ.4.3	Auxiliary components installed	
	Sample probe, reagent probe, microbeads mixer, measuring cell, sipper probe and pre-wash sipper probe installed	Pass
	Waste liner, CC/PC cups, CleanCell, ProCell, PreClean and Assay Cup&Tip trays loaded	Pass
	System prime and system air purge for syringes and reagents	Pass
IQ.4.4	Instrument installation check	
	Temperatures within specifications	Pass
	No alarms during check	Pass
	Air Aspiration Calibration	Pass
	PMT Setting	Pass
	Blank Cell calibration (attached printout)	Pass
	instrument Check (attached printout)	Pass
IQ.4.5	Application installation	
	"Elecsys TSH for Instrument Check" * e-pack does not need TSH application to be downloaded. * GMMI : 0702 8091 200	Pass
	- For Precision Check (OQ.3.3) you need a New TSH e-pack and download TSH (10172) application from cobas link. IC TSH can not be used for Precision Check.	
IQ.4.6	Backup of adjustment data	
	Adjustments data backed up	Pass



*Deviation Report:* Any discrepancies found during the installation must be documented in the space below.

Deviation #1

Investigation

Action taken

Deviation resolved satisfactorily? specify

NA

Deviation #2

Investigation

Action taken

Deviation resolved satisfactorily? specify

NA

Deviation #3

Investigation

Action taken

Deviation resolved satisfactorily? specify

NA



## Installation Qualification for cobasLink

### Description

IQ.5.1	<b>cobasLink connectivity test</b>	
	cobasLink is installed according to the cobas Link cobas pure manual	Pass
	Internet connection is available	Pass
IQ.5.2	<b>cobasLink configuration</b>	
	cobas link Configurator/System Check Latest patches installed	Pass
	General settings are entered according to cobasLink Manual (Laboratory/cobasLink/Utilities incl. certificate )	Pass
	cobas link Configurator/System Check (green traffic light)	Pass
	The CU - configuration was sent to CL and is visible @ cobas Link configurator/Query Tool/ RSi2 > Execute	Pass
IQ.5.3	<b>cobasLink Initiate Upload/ Download</b>	
	The configuration was sent to the TSN server with "Initiate Upload"	Pass
	The application files arrive after "Initiate Download"	Pass
	Perform Sync with cobas link Maintenance>Service>Sync with cobas link Wait 30min	Pass





**Deviation Report:** Any discrepancies found during the installation must be documented in the space below. Roche personnel will then investigate the deviation and decide upon the most appropriate action to be taken.

Deviation #1

Investigation

Action taken

Deviation resolved satisfactorily? specify

NA

Deviation #2

Investigation

Action taken


Deviation resolved satisfactorily? specify

NA



**cobas® pure integrated solutions**

**Operational Qualification**



**cobas®**



## Operational Qualification:

**Notice:**

The steps described in OQ.1 have to be carried out **after a new system installation**, and after any repair action which requires **additional calibration**.

If the service action does not affect the measurement performance, only apply steps OQ.2 and OQ.3 of the Operation Qualification.

**Description**

**OQ.1 Calibration**

Calibration of all photometric parameters successful (attached printout)  yes

Calibration of ISE parameters successful (attached printout)  yes

Calibration of all Immuno parameters successful (attached printout)  yes

Specify the type of calibrator used:

C.F.A.S. lot:

TSH CalSet lot:

**OQ.2 Quality Control**

Specify the type of control used:

PCCC1 lot:

PCCC2 lot:

PCU1 lot:

PCU2 lot:

QC of all photometric parameters within acceptable range (attached printout)  yes

QC of ISE parameters within acceptable range (attached printout)  yes

QC of Immuno parameters within acceptable range (attached printout)  yes

**OQ.3.1 Precision check for ISE**

Perform precision check using PCCC1 n=21

	Number of det.	Expected < CV	Actual CV
Na	21	1,00%	-
K	21	1,20%	-
Cl	21	1,70%	-

} Refer PQ data

Expected precision CV values are only to judge performance of newly installed analyzer, for official specification please refer to assay specific Method sheet.

Precision check for ISE was within acceptable range  yes





QC.3.2 Precision check for Photometric Assays

Perform precision check using PCCC1 n=21

	Number of det.	Expected < CV	Actual CV
ASTP/ASTP2	21	2.00%	-
GLUC3	21	1.00%	-
CREJ2	21	2.50%	-
TP2	21	1.00%	-
CHOL2	21	1.00%	-

} Refer PQ data

Expected precision CV values are only to judge performance of newly installed analyzer, for official specification please refer to assay specific Method sheet.

Precision check for Photometric Assays was within acceptable range

yes

QC.3.3 Precision check for Immunology Assays

Perform precision check using PCU1 n=21 per Channel

New TSH (10172) Reagent

	Number of det.	Expected < CV	Actual CV
TSH Ch.1	21	5.00%	-

} Refer PQ data

check for

yes

QC.3.4 software

results for

yes

for

yes

results for

yes




Deviation #1	
Investigation	
Action taken	NA
Deviation resolved satisfactorily?	specify



**cobas® pure integrated solutions**

**Attachments**



**cobas®**



## Attachments

1. Calibration Results
2. Controls data
3. Precision data (Refer PQ data)

cobas®



## Revision History

*Version*

v.1.0	Date	24-Nov-20	Author	mooniyer
				Initial release
v.2.0	Date	20-Apr-21	Author	mooniyer



# Calibration Monitor

User ID: REDCLIFFE

13/07/2022 17:20

## Calibration Data

13/07/2022 17:20

13/07/2022 17:20:27 (IC)

REDCLIFFE

TSH Unit: µIU/mL

Calib. Lot: 575647

Calibration ID: 0000000001

ASSAY R.P.: 622589/0066887

ProCell: 620137

Result Message: Lot calibration was successfully generated.

Timeout Message:

---Cal 1----- ---Cal 2-----

Target Value

0.000000

1.76

Signal 1

562

25527

Signal 2

569

25723

Diff

--

--

Dup

--

--

System Error

--

--

Calculation Result

--

Calibration Factor

1.00



User ID:

13/07/2022 17:10

Calibration Data

13/07/2022 17:10

13/07/2022 17:10:26 (CC)

CREJ2

Calibration ID: 0000000014

R1 606453/0000623

R2

-----/-----

R3 606453/0000623

STD Name

Lot

First Abs.

Second Abs.

First Initial Abs.

Second Initial Abs.

Data Alarm

--S1---

999999

0.0000

0.0001

0.0376

0.0377

--S2---

539837

0.0263

0.0259

0.2045

0.2046

-Intercept-

- K -

0.0001

14230.8



Calibration Data

13/07/2022 17:10 17:10

13/07/2022 17:10:42 (CC)

GLUC3 Calibration ID: 0000000015

R1 629487/0003430 R2 -----/----- R3 629487/0003430

STD Name	Lot	First Abs.	Second Abs.	First Initial Abs.	Second Initial Abs.	Data Alarm
--S1---	999999	0.0037	0.0035	0.0873	0.0867	
--S2---	539837	0.6512	0.6522	0.7719	0.7718	
-Intercept-	- K -					
	0.0036	16.8184				

**Sample Results Report**

User ID: REDCLIFFE

27/07/2022 14:48

Result Type: First      Sorted by: Date

Sequence No.      Rack ID - Pos.      Sample ID / QC Lot

Comment

Registered

User ID

Q001002	Q30001-1		PCCC1 / 525027	REDCLIFFE	15/07/2022	10:54:59
---------	----------	--	----------------	-----------	------------	----------

**Result Data Alarm**

Test	Sample Type	Dilution	A. U.	Priority	R. P. Lot	R. P. Serial No.	ProCell Lot / Electrode
------	-------------	----------	-------	----------	-----------	------------------	-------------------------

Result Message

ALB2-G	Ser/PI		3.11	CC	Current	R1-R3	619902	0003076
AMYL2	Ser/PI		84.0	CC	Current	R1-R3	618366	0001399
CA2	Ser/PI		8.12	CC	Current	R1-R3	622620	0000354
CHOL2-I	Ser/PI		94.2	CC	Current	R1	640736	0002898
CREJ2	Ser/PI		1.02	CC	Current	R1-R3	606453	0000623
CRP4	Ser/PI		9.58	CC	Current	R1-R3	618926	0008618

# Sample Results Report

User ID: REDCLIFFE

27/07/2022 14:48

Result Type: First Sorted by: Date

Sequence No.	Rack ID - Pos.	Sample ID / QC Lot	Registered
Q001002	Q30001-1	PCCC1 / 525027	REDCLIFFE
			15/07/2022 10:54:59

Test	Result	Data Alarm	A. U.	Priority	R. P. Lot	R. P. Serial No.	ProCell Lot / Electrode
------	--------	------------	-------	----------	-----------	------------------	-------------------------

GGT2-I		55.1					
U/L	Ser/PI		CC	Current	R1-R3	622199	0003410
GLUC3		103					
mg/dL	Ser/PI		CC	Current	R1-R3	629487	0003430
HDLC4		31.0					
mg/dL	Ser/PI		CC	Current	R1-R3	590715	0001401
IRON2		107					
µg/dL	Ser/PI		CC	Current	R1-R3	623943	0002726
LDLC3		57.6					
mg/dL	Ser/PI		CC	Current	R1-R3	572897	0001196
TRIGL		44.2 QCErr					
mg/dL	Ser/PI		CC	Current	R1	618847	0001482
UA2		4.66					
mg/dL	Ser/PI		CC	Current	R1-R3	627564	0003683

System: cobas pure

Serial No.: 2265-09

# Sample Results Report

User ID: REDCLIFFE

27/07/2022 14:49

Result Type: First Sorted by: Date

Sequence No. Rack ID - Pos. Sample ID / QC Lot

Registered

User ID

Q002002 Q30001-2 Name / Lot: PCCC2 / 519198 15/07/2022 10:54:59  
REDCLIFFE

Test Result Data Alarm

Unit Sample Type Dilution A. U. Priority R. P. Lot R. P. Serial No. ProCell Lot / Electrode

Result Message

GGT2-I 228

U/L Ser/PI CC Current R1-R3 622199 0003410

GLUC3 243

mg/dL Ser/PI CC Current R1-R3 629487 0003430

HDLC4 60.5

mg/dL Ser/PI CC Current R1-R3 590715 0001401

IRON2 235

µg/dL Ser/PI CC Current R1-R3 623943 0002726

LDLC3 105

mg/dL Ser/PI CC Current R1-R3 572897 0001196

TRIGL 142 QC Err

mg/dL Ser/PI CC Current R1 618847 0001482

UA2 9.58

mg/dL Ser/PI CC Current R1-R3 627564 0003683

System: cobas pure

Serial No.: 2265-09

**Sample Results Report**

User ID: REDCLIFFE

27/07/2022 14:49

Result Type: First Sorted by: Date

Sequence No.	Rack ID - Pos.	Sample ID / QC Lot	User ID	Registered
Q002002	Q30001-2	PCCC2 / 519198	REDCLIFFE	15/07/2022 10:54:59

Test	Result Data Alarm	A. U.	Priority	R. P. Lot	R. P. Serial No.	ProCell Lot / Electrode
------	-------------------	-------	----------	-----------	------------------	-------------------------

ALB2-G	5.01	CC	Current	R1-R3	619902	0003076
g/dL	Ser/PI					
AMYL2	183	CC	Current	R1-R3	618366	0001399
U/L	Ser/PI					
CA2	12.6	CC	Current	R1-R3	622620	0000354
mg/dL	Ser/PI					
CHOL2-I	171	CC	Current	R1	640736	0002898
mg/dL	Ser/PI					
CREJ2	4.07	CC	Current	R1-R3	606453	0000623
mg/dL	Ser/PI					
CRP4	55.6	CC	Current	R1-R3	618926	0008618
mg/L	Ser/PI					

# Sample Results Report

User ID: REDCLIFFE

13/07/2023 10:30

Result Type: First Sorted by: Date

Sequence No.	Rack ID - Pos.	Sample ID / QC Lot	Registered
Q001083	Q30001-1	PCCC1 / 564978	REDCLIFFE
Comment	Name / Lot:		User ID
			REDCLIFFE
			13/07/2023 09:47:18

Test	Sample Type	Dilution	A. U.	Priority	R. P. Lot	R. P. Serial No.	ProCell Lot / Electrode
------	-------------	----------	-------	----------	-----------	------------------	-------------------------

ALB2-G	Ser/PI	3.35	CC	Current	R1-R3	689517	0000405
g/dL							
ALP2	Ser/PI	111	CC	Current	R1-R3	710401	0006409
U/L							
ALTP	Ser/PI	53.8	CC	Current	R1-R3	704668	0017833
U/L					R2	703735	0005892
APOAT	Ser/PI	95.0	NACL	Current	R1-R3	695028	0000089
mg/dL							
APOBT	Ser/PI	48.4	NACL	Current	R1-R3	631378	0000950
mg/dL							
ASTP	Ser/PI	45.1	CC	Current	R1-R3	700988	0001063
U/L					R2	703735	0005892

System: cobas pure

Serial No.: 2265-09

3 of 16



# Sample Results Report

User ID: REDCLIFFE

13/07/2023 10:30

Result Type: First Sorted by: Date

Sequence No.	Rack ID - Pos.	Sample ID / QC Lot	User ID	Registered
Q001083	Q30001-1	PCCC1 / 564978	REDCLIFFE	13/07/2023 09:47:18

Test	Sample Type	Result	Data	Alarm	A. U.	Priority	R. P. Lot	R. P. Serial No.	ProCell Lot / Electrode
BILD2-D	Ser/PI	0.720	CC	Current	R1-R2	700536	0000883		
BILT3	Ser/PI	0.903	CC	Current	R1-R3	647036	0008342		
CA2	Ser/PI	8.86	CC	Current	R1-R3	690455	0006521		
CHOL2-I	Ser/PI	107	CC	Current	R1	702347	0002265		
CREJ2	Ser/PI	1.00	CC	Current	R1-R3	711464	0004019		
CRP4	Ser/PI	6.05	CC	Current	R1-R3	699752	0003169		
CRP-HS	Ser/PI	7.47	CC	Current	R1-R3	626644	0000861		

System: cobas pure

Serial No.: 2265-09

4 of 16

# Sample Results Report

User ID: REDCLIFFE

13/07/2023 10:30

Result Type: First Sorted by: Date

Sequence No.	Rack ID - Pos.	Sample ID / QC Lot	User ID	Registered
Q001083	Q30001-1	PCCC1 / 564978	REDCLIFFE	13/07/2023 09:47:18

Test	Sample Type	Dilution	A. U.	Priority	R. P. Lot	R. P. Serial No.	ProCell Lot / Electrode
------	-------------	----------	-------	----------	-----------	------------------	-------------------------

GGT2-I U/L	Ser/PI	52.7	CC	Current	R1-R3	711304	0017237
GLUC3 mg/dL	Ser/PI	105	CC	Current	R1-R3	702348	0005174
HDLC4 mg/dL	Ser/PI	31.4	CC	Current	R1-R3	704173	0000729
IRON2 µg/dL	Ser/PI	109	CC	Current	R1-R3	711465	0002092
ISE CL mmol/L	Ser/PI	83.1OBS.EL	ISE	Current	IS	669181	0004406
					DIL	709792	0013894
					REF	668700	0011367
ISE K mmol/L	Ser/PI	3.49OBS.EL	ISE	Current	IS	669181	0004406
					DIL	709792	0013894
					REF	668700	0011367

# Sample Results Report

User ID: REDCLIFFE

13/07/2023 10:30

Result Type: First Sorted by: Date

Sequence No.	Rack ID - Pos.	Sample ID / QC Lot	User ID	Registered
Q001083	Q30001-1	PCCC1 / 564978	REDCLIFFE	13/07/2023 09:47:18

Test	Sample Type	Dilution	A. U.	Priority	R. P. Lot	R. P. Serial No.	ProCell Lot / Electrode
------	-------------	----------	-------	----------	-----------	------------------	-------------------------

ISE NA		112.6 OBS.EL	ISE	Current	669181	0004406	G2899 A1058
mmol/L	Ser/PI				709792	0013894	
					668700	0011367	

LDHI2		169 ReagEx OBS.RR	CC	Current	618368	0001335	
U/L	Ser/PI						
LDLC3		59.6	CC	Current	619975	0003614	
mg/dL	Ser/PI						
PHOS2		3.58	CC	Current	703308	0006906	
mg/dL	Ser/PI						
TP2		5.03	CC	Current	713203	0008892	
g/dL	Ser/PI						
TRIGL		114	CC	Current	703177	0000243	
mg/dL	Ser/PI						

System: cobas pure

Serial No.: 2265-09

6 of 16

# Sample Results Report

User ID: REDCLIFFE

13/07/2023 10:30

Result Type: First Sorted by: Date

Sequence No.	Rack ID - Pos.	Sample ID / QC Lot	User ID	Registered
--------------	----------------	--------------------	---------	------------

Q001083 Q30001-1 PCCC1 / 564978 REDCLIFFE 13/07/2023 09:47:18

Test	Sample Type	Dilution	A. U.	Priority	R. P. Lot	R. P. Serial No.	ProCell Lot / Electrode
------	-------------	----------	-------	----------	-----------	------------------	-------------------------

UA2	Ser/PI	4.56	CC	Current	R1-R3	695529	0002745
UIBC-I	Ser/PI	224	CC	Current	R1-R3	694806	0001260
UREAL	Ser/PI	37.6	CC	Current	R1-R3	712545	0022867

Result Message

# Sample Results Report

User ID: REDCLIFFE

13/07/2023 10:30

Result Type: First      Sorted by:      Date

Sequence No.	Rack ID - Pos.	Sample ID / QC Lot	User ID	Registered
Q002083	Q30001-2	PCCC2 / 595393	REDCLIFFE	13/07/2023 09:47:18

Name / Lot: PCCC2 / 595393

Test	Sample Type	Dilution	A. U.	Priority	R. P. Lot	R. P. Serial No.	ProCell Lot / Electrode
------	-------------	----------	-------	----------	-----------	------------------	-------------------------

ALB2-G			5.30				
g/dL	Ser/PI			Current	689517	0000405	
ALP2			258				
U/L	Ser/PI			Current	710401	0006409	
ALTP			132				
U/L	Ser/PI			Current	704668	0017833	
					703735	0005892	
ASTP			144				
U/L	Ser/PI			Current	700988	0001063	
					703735	0005892	
BILD2-D			2.06				
mg/dL	Ser/PI			Current	700536	0000883	
BILT3			3.42				
mg/dL	Ser/PI			Current	647036	0008342	

# Sample Results Report

User ID: REDCLIFFE

13/07/2023 10:30

Result Type: First      Sorted by:      Date

Sequence No.	Rack ID - Pos.	Sample ID / QC Lot	User ID	Registered
--------------	----------------	--------------------	---------	------------

Q002083      Q30001-2      Name / Lot: PCCC2 / 595393      REDCLIFFE      13/07/2023      09:47:18

Test	Sample Type	Dilution	A. U.	Priority	R. P. Lot	R. P. Serial No.	ProCell Lot / Electrode
------	-------------	----------	-------	----------	-----------	------------------	-------------------------

CA2	Ser/PI	13.6	CC	Current	R1-R3	690455	0006521
CHOL2-I	Ser/PI	170	CC	Current	R1	702347	0002265
CREJ2	Ser/PI	3.70	CC	Current	R1-R3	711464	0004019
CRP4	Ser/PI	51.7	CC	Current	R1-R3	699752	0003169
GGT2-I	Ser/PI	225	CC	Current	R1-R3	711304	0017237
GLUC3	Ser/PI	247	CC	Current	R1-R3	702348	0005174
HDLC4	Ser/PI	52.2	CC	Current	R1-R3	704173	0000729

Result Message

# Sample Results Report

User ID: REDCLIFFE

13/07/2023 10:30

Result Type: First Sorted by: Date

Sequence No.	Rack ID - Pos.	Sample ID / QC Lot	Registered
Q002083	Q30001-2	PCCC2 / 595393	REDCLIFFE
Name / Lot:		13/07/2023	09:47:18

Test	Sample Type	Dilution	A. U.	Priority	R. P. Lot	R. P. Serial No.	ProCell Lot / Electrode
------	-------------	----------	-------	----------	-----------	------------------	-------------------------

IRON2	Ser/PI	244	CC	Current	R1 R3	0002092	
ISE CL	Ser/PI	101.0OBS.EL	ISE	Current	IS	0004406	Y9695 A1058
					DIL	0013894	
					REF	0011367	
ISE K	Ser/PI	7.06OBS.EL	ISE	Current	IS	0004406	P8015 A1058
					DIL	0013894	
					REF	0011367	
ISE NA	Ser/PI	136.0OBS.EL	ISE	Current	IS	0004406	G2899 A1058
					DIL	0013894	
					REF	0011367	
PHOS2	Ser/PI	8.01	CC	Current	R1-R3	0006906	

System: cobas pure

Serial No.: 2265-09

10 of 16

# Sample Results Report

User ID: REDCLIFFE

13/07/2023 10:30

Result Type: First      Sorted by:      Date

Sequence No.	Rack ID - Pos.	Sample ID / QC Lot	Registered
Q002083	Q30001 2	PCCC2 / 595393	REDCLIFFE
Comment	Name / Lot:		User ID
			REDCLIFFE
			13/07/2023 09:47:18

Test	Sample Type	Dilution	A. U.	Priority	R. P. Lot	R. P. Serial No.	ProCell Lot / Electrode
------	-------------	----------	-------	----------	-----------	------------------	-------------------------

TP2							
g/dL	Ser/PI		CC	Current	R1-R3	713203	0008892
TRIGL		212					
mg/dL	Ser/PI		CC	Current	R1	703177	0000243
UA2		9.52					
mg/dL	Ser/PI		CC	Current	R1-R3	695529	0002745
UIBC-I		260					
µg/dL	Ser/PI		CC	Current	R1-R3	694806	0001260
UREAL		116					
mg/dL	Ser/PI		CC	Current	R1-R3	712545	0022867

Result Message

