


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 230629A/N-120	Page	: 1 of 1
CALIBRATED ON.	: 04 / 07 / 2023	Ref. Std.	: IS:12508:1988
CALIBRATION DUE.	: 03 / 07 / 2024	C.P. No.	: ACS/CP/10
SRF No. & DATE	: 230629A/N Dt. 29.06.2023	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342723000000684F Dt. 06.07.2023	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. 2nd Floor, 1-4-168/4/4/NR, Satya Complex, Loyola College Road, Landmark Karur Vysya Bank, Old Alwal, Secunderabad, Hyderabad – 500010		
ITEM CALIBRATED	: Centrifuge	Make	: Remi
		Range	: (0 to 6000) RPM
		L. Count	: ID No. : RCL/BIO/Cent-02
		Location	: S. No. : ZIBN-08851
ENVIRONMENT CONDITION:	Temp. { 25°C + 3°C} R.H. {40% to 60%}		
STANDARD INSTRUMENTS USED :			
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	C-220912-6-1
			DUE DATE
			12/09/2023
RESULTS	: Mechanical Discipline		
* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)	
1000	1001.50	- 1.50	
2000	2002.50	- 2.50	
4000	4005.74	- 5.74	
6000	6006.76	- 6.76	
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 2.50 rpm of reading at coverage Factor $k=1.96$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

Auth. Signatory
S.K. Sharma
CEO/TM
C-23, Sec-10, Noida