



# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b> : 230629A/N-122	<b>Page</b> : 1 of 1
<b>CALIBRATED ON.</b> : 04 / 07 / 2023	<b>Ref. Std.</b> : CG-11
<b>CALIBRATION DUE.</b> : 03 / 07 / 2024	<b>C.P. No.</b> : ACS/CP/07
<b>SRF No. &amp; DATE</b> : 230629A/N Dt. 29.06.2023	<b>Phy. Inspection</b> : O.K.
<b>ULR No. &amp; ISSUE DATE</b> : CC342723000000686F Dt. 06.07.2023	<b>Cal. At</b> : Site / Lab

<b>CALIBRATED FOR</b> : M/s Redcliffe Lifetech Pvt. Ltd. 2nd Floor, 1-4-168/4/NR, Satya Complex, Loyola College Road, Landmark Karur Vysya Bank, Old Alwal, Secunderabad, Hyderabad – 500010
---

<b>ITEM CALIBRATED</b> : Digital Temperature Indicator	
Make : RoHS	Range : (-50 to 110) °C
L. Count :	ID No. : RCL/BIO/D.Temp-01
Location : Clinical Biochemistry	S. No. : N.A

**ENVIRONMENT CONDITION:** Temp. { 25°C ± 3°C} R.H. {40% to 60%}

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2022/0846	10/11/2023

RESULTS : ET Discipline		
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)
2.04	2	0.04
4.08	4	0.08
6.12	6	0.12
8.18	8	0.18
(average of readings)		

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 1.35°C at coverage Factor k=1.96.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmy statement not provided since not required.

Prepared by

....End of Calibration Certificate....

