

ADVANCE CALIBRATION SERVICES


(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



CC-3427



Calibration Certificate

CERTIFICATE NO.	: 230629A/N-128	Page	: 1 of 1	
CALIBRATED ON.	: 04 / 07 / 2023	Ref. Std.	: CG-11	
CALIBRATION DUE.	: 03 / 07 / 2024	C.P. No.	: ACS/CP/07	
SRF No. & DATE	: 230629A/N Dt. 29.06.2023	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC342723000000692F Dt. 06.07.2023	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. 2nd Floor, 1-4-168/4/4/NR, Satya Complex, Loyola College Road, Landmark Karur Vysya Bank, Old Alwal, Secunderabad, Hyderabad – 500010			
ITEM CALIBRATED	: Digital Temperature Indicator			
	Make : RoHS	Range	: (-50 to 110) °C	
	L. Count :	ID No.	: RCL/HEM/D.Temp-07	
	Location : Haematology	S. No.	: N.A	
ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C } R.H. {40% to 60%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2022/0846	10/11/2023
RESULTS	: ET Discipline			
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)		
2.16	2	0.16		
4.10	4	0.10		
6.20	6	0.20		
8.22	8	0.22		
(average of readings)				

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 1.35°C at coverage Factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

