


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO. : 230629A/N-114	Page : 1 of 1
CALIBRATED ON. : 29 / 06 / 2023	Ref. Std. : DKD R-5-7
CALIBRATION DUE. : 28 / 06 / 2024	C.P. No. : ACS/CP/11
SRF No. & DATE : 230629A/N Dt. 29.06.2023	Phy. Inspection : O.K.
ULR No. & ISSUE DATE : CC342723000000678F Dt. 06.07.2023	Cal. At : Site / Lab
CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd. 2nd Floor, 1-4-168/4/4/NR, Satya Complex, Loyola College Road, Landmark Karur Vysya Bank, Old Alwal, Secunderabad, Hyderabad - 500010	
ITEM CALIBRATED : Digital Thermo Hygrometer	
Make : Digital Series	Range : (0 to 50)°C, (10-99)%
L. Count :	ID No. : RCL/SR/HGM-03
Location : Sample Receiving Dept.	S. No. : N.A
ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C } R.H. { 40% to 60% }	

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
RH & Temp. Sensor with Indicator	ACS/RHTIS/01/ 20365872	DVG	CC210122000001272F	16/07/2023

RESULTS : Thermal Discipline			
* U. C. INSTT. READS	STANDARD READING		ERROR
(°C)	15.46	14.95	0.51
	25.44	24.93	0.51
	40.50	39.96	0.54
(average of readings)			
(%)	30.74	29.96	0.78
	60.72	59.94	0.78
	90.62	89.94	0.68
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 2.1°C, 2.5% at coverage Factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

...End of Calibration Certificate...

