

## ADVANCE CALIBRATION SERVIO

(ISO: 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob.: +91-9891121066, 9868309579 • E-mail: advance.cal@gmail.com





## Calibration Certificate

CERTIFICATE NO.	: 230629/	: 230629A/N-102			Page	- :1	of 1	
CALIBRATED ON. : 29 / 06 /		2023			Ref. Std.	: 19	SO:8655:6	
CALIBRATION DUE. : 28		28 / 06 / 2024			C.P. No.			
SRF No. & DATE : 230629A/N Dt. 29.06.2023					Phy. Inspection : O.K.			
ULR No. & ISSUE DATE : CC342723000000666F Dt. 06.07.2023					Cal. At	Cal. At : Site / Lab		
CALIBRATED FOR	2nd Lan	cliffe Lifetech Floor, 1-4-10 dmark Karur lerabad – 500	6 <mark>8/4/4/NR,</mark> Vysya Bar	Satya Co nk, Old Al	omplex,Loyola Co wal, Secunderab	ollege ad,	Road,	
ITEM CALIBRATE	: Micro Pi	oette						
Make : Fine Care Corporation					Range : (10 to 100) µl			
L. Count :					ID No. : RCL/HEM/PIP-02			
	Location : Haematology					S. No. : PL572152		
ENVIRONMENT CO	ONDITION: Te	emp. { 25°C -	- 3°C} F	R.H. {40%	to 60%}			
STANDARD INSTR	UMENTS USED	:						
INSTT. NAME	ID No. / S. No.		TRACEABILITY		CERTIFICATE No.		DUE DATE	
Weighing Balance	ACS/DB/02/ D449927064		BlueBox		BBT/398/JAN/23		05/01/2024	
RESULTS :	Mechanical	Discipline						
Nominal Capacity (µI)	Graduation (µl)	Nominal Value (µl)		Volume at 20 °C (µI)		Error (µl)		
	10			10.10			(-) 0.10	
100		50			50.19		(-) 0 19	

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is  $\pm$  0.1  $\mu$ l at coverage factor k=2.

100.40

Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.

The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.

100

The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.

Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use

Confirmity statement not provided since not required.





(-) 0.40