


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b> : 230629A/N-103	<b>Page</b> : 1 of 1
<b>CALIBRATED ON.</b> : 29 / 06 / 2023	<b>Ref. Std.</b> : ISO:8655:6
<b>CALIBRATION DUE.</b> : 28 / 06 / 2024	<b>C.P. No.</b> : ACS/CP/08
<b>SRF No. &amp; DATE</b> : 230629A/N Dt. 29.06.2023	<b>Phy. Inspection</b> : O.K.
<b>ULR No. &amp; ISSUE DATE</b> : CC342723000000667F Dt. 06.07.2023	<b>Cal. At</b> : Site / Lab

**CALIBRATED FOR** : M/s Redcliffe Lifetech Pvt. Ltd.  
2nd Floor, 1-4-168/4/4/NR, Satya Complex, Loyola College Road,  
Landmark Karur Vysya Bank, Old Alwal, Secunderabad,  
Hyderabad – 500010

<b>ITEM CALIBRATED</b> : Micro Pipette	
<b>Make</b> : Fine Care Corporation	<b>Range</b> : (100 to 1000) µl
<b>L. Count</b> :	<b>ID No.</b> : RCL/HEM/PIP-03
<b>Location</b> : Haematology	<b>S. No.</b> : QG602883

**ENVIRONMENT CONDITION:** Temp. { 25°C ± 3°C } R.H. { 40% to 60% }

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/398/JAN/23	05/01/2024

RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		100	100.36	(-) 0.36
1000		500	501.67	(-) 1.67
		1000	1003.31	(-) 3.31

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 0.6 µl at coverage factor k=2.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

