



Ref No. TS/6618/SBP

Date :09 Jan '2023

**PathKind Diagnostics Private Limited
C/o Sharda Narayan Hospital
Sahadatpuram, Mau (U.P)
275101**

Dear Sir,

As desired by you please find the Calibration report

1. Year of manufacturing : 2017 Easylyte Sr No. 59845CNKC

2. Date of Calibration -09th Jan'2023

Test Voltages / Test Temperature -

Test Voltage	Actual Voltage	OK/not OK
Logic 5 V	5.06V	OK
Line voltage	230V	OK
Logic 12V	12V	OK
Calibration Value (Na)	59.45	OK



Calibration Value (K)	59.71	OK
Calibration Value (Cl)	47.21	OK
Electrode Values(Na)	131	OK
Electrode Value(K)	85	OK
Electrode Value(Cl)	57	OK
Pump calibration		OK
Fluid Test		OK

.....2

Instrument checked with Medica Controls

Thanking you and assuring of our prompt attention at all times.

Yours faithfully,

For TRANSASIA BIO-MEDICALS LTD.



