



Calibration Certificate

Instrument Name : Urometer120
Serial Number : 12076171

Customer Name : RLM ALL CARE DIAGNOSTICS

Address : Bandlaguda,Chandrayanagutta,Hyderabad-500005
Standard used : Self Calibration Mode
Standard Traceability : In House Certificate

Calibration

The instrument function well in all the systems check, self check and maintenance Procedures. The equipment was calibration and the Red, Blue & Green (R,B & G) filters of all the parameters and top & bottom were found to be within the acceptable range of 333 ± 10 .

QC Performance : Acceptable
Self Calibration : Acceptable

Performance Qualification : Passed

Calibration Date : 05.06.2023

Calibration : **Passed**

Calibration Valid up to :04.06.2024

For Abbott Diagnostics Medical Pvt Ltd.

Upendrakumar V
Engineer Technical Services
M: 9052723178

