

# PROFICIENCY TESTING REPORT



ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
NABL accredited program as per ISO/IEC 17043:2010 standard Organized By Department of Hematology, AIIMS, New Delhi-110029

Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

**EQAP CODE No.** : 2854 **Distribution No.:** 159-G Month/Year: March/2023

**Instrument ID:** 109YAXH03477

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,

Tel: 9013085730, E-Mail: accuracy2000@gmail.com Date of issue & status of the report: 01-05-2023[Final].

# **CBC** and Retic Assessment

		Among Lab (Accuracy Testing)			ıg)	Within Lab (Precision Testing)					
Test Parameters	S.No.	Your Result 1		Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty		Results		Uncertainty of Assigned Values	Z Score
WBC x10³/μl	1	0.9	0.7	1.6	8.21	0.046	-5.94	0.2	0.1	0.007	0.90
RBC x10 <sup>6</sup> /μl	1	4.39	4.39	8.78	8.8	0.010	-0.10	0	0.04	0.003	-1.08
Hb g/dl	1	14	13.9	27.9	25.9	0.024	3.17	0.1	0.1	0.008	0.00
НСТ%	1	40.1	40	80.1	80.5	0.180	-0.09	0.1	0.3	0.027	-0.54
MCV-fl	1	91.2	91	182.2	183.65	0.290	-0.19	0.2	0.3	0.025	-0.27
MCH-Pg	1	31.8	31.6	63.4	59	0.073	2.58	0.2	0.2	0.017	0.00
MCHC-g/dl	1	34.9	34.7	69.6	64.05	0.136	1.62	0.2	0.3	0.020	-0.34
Plt. x10³/μl	1	185	185	370	293	1.455	2.08	0	4.5	0.344	-0.87
Retic %	2	2.6	2.5	5.1	8.15	0.186	-0.65	0.1	0.4	0.026	-1.16

# P.S. Assesment

		YOUR REPORT	CONSENSUS REPORT		
Nrbcs=22 , Poly=49 L=47, E=02,		Mono/Promono=01 , B1=00 P.M.=00,	Poly: 55-66, Lympho: 24-34, Mono: 1-4, Eosino: 1-3, blast/Promyelo/Myelo/Meta: 0-5		
		Anisocytosis +, Poiilocytosis Mild, Tear	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, Microcytosis, Hypochromia; Mild: Poikilocytosis, Target cells, Sickle shaped cells, tear drop cells		
Diagnosis	3	Thalassemia	Thalassemia/Haemoglobinopathy		

### **COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S No	Total participants	Total No.	C 0 2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
Test parameters	5.NU.	current dist. 159G	responded	Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10³/μl	1	248	248	<mark>8</mark> 7.5	88.31	4.44	3.63	8.06	8.06
RBC x10 <sup>6</sup> /μl	1	248	248	79.84	87.9	10.48	6.05	9.68	6.05
Hb g/dl	1	248	248	88.71	92.34	4.84	3.63	6.45	4.03
HCT%	1	248	2 <mark>48</mark>	94.35	89.92	4.44	4.84	1.21	5.24
MCV-fl	1	248	248	96.77	91.53	2.42	3.23	0.81	5.24
MCH-Pg	1	248	247	91.9	94.74	6.07	0.4	2.03	4.86
MCHC-g/dl	1	248	248	94.76	90.73	4.84	2.82	0.4	6.45
Plt. x10³/μl	1	248	248	92.74	93.95	5.24	4.03	2.02	2.02
ReticCount%	2	248	216	91.2	84.26	5.56	9.26	3.24	6.48
PS Assessment	3	248	219	Satisfactory	:90.74%, Bo	orderline Sat	:8.06%, Uı	nsatisfactory	:1.20%

#### \*Comments:

1). Among Lab (EQA): CBC result for WBC & HB unacceptable, please check calibration/human error.Remaining results acceptable.

2). Within Lab (IQA): Precision acceptable.

**Note-1: EQA** (External Quality Assurance): Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** (Internal Quality Assurance): Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to  $\pm 2$ : Acceptable, Z score  $\pm 2$  to  $\pm 3$ : Warning Signal, Z score  $> \pm 3$ : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample  $(\bar{x}-\bar{y})$  should be smaller than the check value (0.3\*SDPA).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

**Note 10:** Reports are kept confidential.

Report authorized by,

Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----

Title	PT/ EQAS EVALUATION RECORD
Document Number	FRM.QCM.03
Version	02
Amendment No	00
Effective Date	02.06.2023



Date	e of Investigation: 17 0 7 23
PT	/EQAS Set Identification: 159- G
Da	te of PT/EQAS: March-23
Ac	ceptable/ Unacceptable Results - WBC, Hb, MCH, PIE.
Ac	ceptable Result Range:
Pr	evious Trends/ Unacceptable Results from this Analyte/ Test:
	No Trend.
□	Transcription error (may be pre- or post-analytical factors) Wrong method has been registered for analysis or method change not updated.  etails of Investigation:
	ethodological Instrument function checks (e.g., temperatures, blank readings, pressures) not performed as necessary, or results not within acceptable range.  Scheduled instrument maintenance not performed appropriately.
	Incorrect instrument calibration.
	Standards or reagents improperly reconstituted and stored, or inadvertently used beyond expiration date.
	Instrument probes misaligned.
	Problem with instrument data processing functions. The laboratory may need to contact the manufacturer to
	evaluate such problems.
	Problem in manufacture of reagents / standards, or with instrument settings specified by manufacturer
	Carry-over from previous specimen.
	Automatic pipettor not calibrated to acceptable precision and accuracy.
	Imprecision from result being close to detection limit of method.
	QC material not run within expiration date, or improperly stored.

Lupin Diagnostics (Lupin Diagnostics Limited)	Page 1 of 4
Site: All Locations	CONFIDENTIAL: Authorized for internal use only

Title	PT/ EQAS EVALUATION RECORD
Document Number	FRM.QCM.03
Version	02
Amendment No	00
Effective Date	02.06.2023

Lupin Diagnostics (Lupin Diagnostics Limited)

Site: All Locations



Page 2 of 4

CONFIDENTIAL: Authorized for internal use only


	QC material not run at relevant analyte concentration.								
_	Result not within reportable range (linearity) for instrument / reagent system.								
	Obstruction of instrument tubing / orifice by clot or protein.								
	Incorrect incubation times.								
De	tails of Investigation:								
Te	chnical								
	EQA material improperly reconstituted.								
	Testing delayed after reconstitution of EQA material (with problem from evaporation or deterioration).								
	Sample not placed in proper order on instrument.								
	Result released despite unacceptable QC data.								
	QC data within acceptable limits but showed trend suggestive of problem with the assay.								
	Inappropriate quality control limits / rules. If the acceptable QC range is too wide, the probability increases that a result will fall within the acceptable QC range yet exceed acceptable limits for EQA.								
	Manual pipetting / diluting performed inaccurately, at an incorrect temperature or with incorrect diluent.								
	Calculation error or result reported using too few significant digits.								
	Secondary specimen tubes incorrectly labeled.								
	In addition to above discipline specific errors may also occur								
D -	etails of Investigation:								
P	roblem with PT/EQAS Material								
	Matrix effects: The performance of some instrument / method combinations may be affected by the matrix of the PT/EQAS sample. This can be overcome to some extent by assessing participants in peer groups – to be done by the PT/EQAS provider.								
	Non-homogenous test material due to variability infill volumes, inadequate mixing, or inconsistent heating of								
	lyophilized specimens.								
	Non-viable samples for microbiology PT/EQAS program.								
	Haemolysis on an immune-haemtology program samples.								
_	Details of Investigation:								

Title	PT/ EQAS EVALUATION RECORD
Document Number	FRM.QCM.03
Version	02
Amendment No	00
Effective Date	02.06.2023



-							
⇒ro	blem with PT/EQAS Evaluation						
Peer group not appropriate.							
3	Inappropriate target value: Target values developed from participant consensus can be inappropriate from non-homogeneous testing material or lingering ("masked") outliers. However, occasional inappropriate to values occur in every PT program. Inappropriate evaluation interval: An evaluation interval may be inappropriately narrow e.g. if ± 2 standard deviation units are used with an extremely precise method the acceptable range may be much narrower than needed for clinical usefulness.						
	Incorrect data entry by PT provider.						
De	rails of Investigation:						
No	Explanation: Attributed to Random Error						
Su							
100	mmary of Investigation:  ABC howceptable performance due to method a detection as  neution in previous RCA Report.  There parameter the performance found satisfactory.						
	mmary of Investigation:  LIBC how ceptable performance due do method a detection as  mention in previous RCA Report.  There parameter performance found satisfactory.  There parameter according to the parameter of the parameter						
	as patient data affected? & Corrective action taken if Patient data was affected.						
	as patient data affected? & Corrective action taken if Patient data was affected.						
W	as patient data affected? & Corrective action taken if Patient data was affected.						
W	as patient data affected? & Corrective action taken if Patient data was affected.						
W	as patient data affected? & Corrective action taken if Patient data was affected.						

Lupin Diagnostics (Lupin Diagnostics Limited)	Page 3 of 4
Site: All Locations	CONFIDENTIAL: Authorized for internal use only





Conclusions					1 . 1 . 1	Manuscania	610
Balld	ON	above	fundings	cone	maeg	anace spra	ble.
Conclusions Based performa	nop	due to	Fransition	al a	sample	perform	ante
Perton	s C.C.c			P	1	) 0	
nconitor	clo	sely un	next cycl	e.			

Quality Manager/ Team Leader

Date: 17/07/23

Lab Head

LUFIN ELACHOSTIC

Date: 17/07/23

Behind Zopadi Cantin,
Opp. Monica D.Ed. College
Savedi, Ahmednagar-414001

Lupin Diagnostics (Lupin Diagnostics Limited)	Page 4 of 4
Site: All Locations	CONFIDENTIAL: Authorized for internal use only



Reference Laboratory- Lupin Diagnostics, NRL Date of study conducted- 13.07.2023

Sr No				Sample-1	
31 140	Parameters		Mrs. Kamal Takale		
1	DDG.	NRL	HLM Varad	%Diff	Reference range
1	RBC	3.23	3.29	-1.82	3.8-6
3	HB	9.50	9.60	-1.04	11.5-17
	PCV	28.30	28.70	-1.39	35-52
4	MCV	87.50	87.10	0.46	76-100
5	MCH	29.30	29.00	1.03	27-34
6	MCHC	33.50	33.40	0.30	32-35
7	RDWCV	21.00	20.90	0.48	11.0-17
8	RDWSD	63.80	55.40	15.16	37-49
9	PLT	782.00	606.00	29.04	150-400
10	PCT	0.76	0.50	52.00	0.15-0.40
11	MPV	9.80	8.30	18.07	8.0-11
12	PDW	18.00	13.10	37.40	11.0-22
13	PLCC	267.00	131.00	103.82	44-140
14	PLCR	34.20	21.60	58.33	18-50
15	WBC	142.87	170.90	-16.40	3.5-10
16	NEUT	69.86	94.63	-26.18	1.6-7
17	LYMP	18.62	20.77	-10.35	1.0-3
18	MONO	7.79	8.66	-10.05	0.2-0.8
19	EOS	4.16	7.89	-47.28	0.0-0.50
20	BASO	0.00	0.00	0.00	0.0-0.30
21	LIC	42.44	38.95	8.96	0.0-0.13
22	NEUT%	69.60	71.70	-2.93	40-73
23	LYM%	18.50	15.70	17.83	15-45
24	MONO%	7.80	6.60	18.18	4.0-12
25	EOS%	4.10	6.00	-31.67	0.5-7
26	BASO%	0.00	0.00	0.00	0.0-2.0
27	LIC	42.30	29.50	43.39	0.0-0.10

		the Managaments of		Sample-2	
Sr No	Parameters		Mr. Prabhakar G		P 40 CL 40 C C C C C C C C C C C C C C C C C C
	A Comment of the Comm	NRL	HLM Varad	%Diff	Reference range
1	RBC	2.61	2.69	-2.97	3.8-6
2	НВ	8.80	8.80	0.00	11.5-17
3	PCV	25.30	26.50	-4.53	35-52
4	MCV	96.90	98.50	-1.62	76-100
5	НСН	33.60	32.70	2.75	27-34
6	HCHC	34.70	33.20	4.52	32-35
7	RDWCV	23.20	23.60	-1.69	11.0-17
8	RDWSD	73.90	68.00	8,68	37-49
9	PLT	17.00	33.00	-48.48	150-400
10	PCT	0.01	0.03	-66.67	0.15-0.40
11	MPV	7.80	8.50	-8.24	8.0-11
12	PDW	14.50	9.60	51.04	11.0-22
13	PLCC	4.00	8.00	-50.00	44-140
14	PLCR	20.90	24.90	-16.06	18-50
15	WBC	5.93	9.15	-35.19	3.5-10
16	NEUT	1.97	8.14	-75.80	1.6-7

Lupin Diagnostics, (HLM Varad)	
Inter laboratory comparison study report of	Complete Blood count



17	LYMP	0.27	0.32	-15.63	1.0-3
18	MONO	1.58	0.55	187.27	0.2-0.8
19	EOS	0.01	0.00	0.00	0.0-0.50
20	BASO	0.01	0.01	0.00	0.0-0.15
21	LIC	2.09	0.13	1507.69	0.0-0.10
22	NEUT%	51.50	90.20	-42.90	40-73
23	LYM%	6.90	3.60	91.67	15-45
24	MONO%	41.10	6.10	573.77	4.0-12
25	EOS%	0.30	0.00	0.00	0.5-7
26	BASO%	0.20	0.10	100.00	0.0-2.0
27	LIC	54.40	1.50	3526.67	0.0-0.10

# Observations-

- ✓ >80% Clinical correlation noted in both samples.
- High % Difference noted due to statistical limitations.

## Conclusion:

Based on obtained result recovery Inter laboratory comparison study successfully passed for CBC test parameter.

Documented By (Mr. Mahesh B)

Approved by (Dr. Sagar K)

#### Results Run Date 12/07/2023 01:07:07 PM Operator ABX Sample ID HA00342684 Last Name Rack/Pos First Name Department Age Gender Physician Patient ID Type Standard Birth Date Sample comments Recommended actions Range Slide review 3.80 - 6.00 10%/µL 2.69 RBC 11.5 - 17.0 WBC g/dL HGB 8.8 L RBC LYM Interference 35.0 - 52.0 % 26.5 HCT Abnormal Differentiation 76.0 - 100.0 Susp. Pathologies 98.5 µm³ MCV 27.0 - 34.0 Anemia 32.7 pg MCH Anisocytosis 32.0 - 35.0 g/dL 33.2 Thrombopenia MCHC 11.0 - 17.0 % PLT aggregate ? RDW-CV 23.6 Lymphopenia 37.0 - 49.0 µm³ 68.0 H RDW-SD Neutrophilia Range PLT Left shift Malaria P. falciparum ? 150 - 400 103/µL 33 L\* PLT % 0.15 - 0.40 0.03 1\* PCT µm<sup>3</sup> 8.0 - 11.0 8.5 MPV 11.0 - 22.0 μm³ 9.6 1\* PDW 44 - 140 103/µL 8 P-LCC 20 18.0 - 50.0 24.9 % P-LCR DIF Range 3.50 - 10.00 $10^3/\mu L$ 9.15 WBC Range % Range 40.0 - 73.0 h\* 1.60 - 7.00 90.2 H\* 8.14 NEU 15.0 - 45.0 1\* 1.00 - 3.00 3.6 0.32 L\* LYM 4.0 - 12.0 \* 6.1 0.20 - 0.80 0.55 MON 0.5 - 7.0 1\* 0.00 - 0.50 0.0 0.00 EOS 0.0 - 2.0 \* 0.00 - 0.15 0.1 0.01 BAS 0.0 - 1.0 h\* 0.00 - 0.10 h\* LIC

# Slide Review

	Myeloblast	Anisocytosis
Neutrophil	The state of the s	Hypochromia
Lymphocyte	Promyelocyte	Polychromasia
Monocyte	Myelocyte	Poikilocytosis
Eosinophil	Metamyelocyte	Microcytosis
Basophil	Blast	Macrocytosis
Atypical Lymphocyte	Target Cell	Platelet Clumps
Other	Sickle Cell	
Reviewed on	by	Signature :

# ILC - MPL Result

#### Results: Operator LUPIN Run Date 07/13/2023 06:57:51 PM Sample ID HA00342684 Last Name Prabhatan. Gunde Rack/Pos 041731/1 First Name Department Age Gender physician Patient ID Type Standard Birth Date Sample comments Recommended actions Slide review Range Aletinis 3.80 - 6.00 10%/HL RBC 2.61 WBC g/dL 11.5 - 17.0 HGB 8.8 RBC LYM Interference 35.0 - 52.0 MON Interference 0% 25.3 HCT Abnormal Differentiation 76.0 - 100.0 mu<sub>3</sub> MCV 96.9 PLT 27.0 - 34.0 33.6 pg MCH RBC PLT Interference g/dL 32.0 - 35.0 34.7 Susp. Pathologies MCHC 11.0 - 17.0 96 23.2 Anemia RDW-CV 100 Anisocytosis 37.0 - 49.0 um<sup>3</sup> 73.9 14 RDW-5D Thrombopenia Range PLT aggregate 7 103/UL 150 - 400 Lymphopenia 17 PLT 0.15 - 0.4096 Monocytosis 0.01 PCT Large Immature Cells 8.0 - 11.0 $\mu m^2$ 7.8 1\* MPV Left shift 11.0 - 22.0 µm³ 14.5 PDW 44 - 140 4 1\* 103/µL P-LCC 20 20 18.0 - 50.0 20.9 % P-LCR DIF Range 3.50 - 10.00 $10^3/\mu L$ 5.93 WBC Range 96 # Range 40.0 - 73.0 1.60 - 7.00 51.5 1.97 NEU 15.0 - 45.0 1.00 - 3.00 6.9 1\* 0.27 L\* LYM 0.20 - 0.80 41.1 his 4.0 - 12.01.58 MON 0.5 - 7.00.3 1\* 0.00 - 0.50 0.01 EO5 0.0 - 2.00.00 - 0.15 0.2 \* 0.01 BAS 0.0 - 1.00.00 - 0.10 54.4 HW 2.09 Slide Review

Neutrophil	Myeloblast	Anisocytosis
Lymphocyte	Promyelocyte	Hypochromia
Monocyte	Myelocyte	Polychromasia
Eoslnophil	Metamyelocyte	Paikllocytosis
Basophil	Blast	Microcytosis
Atypical Lymphocyte	Target Cell	Macrocytosis
Other	Sickle Cell	Platelet Clumps
Reviewed on	by	Signature :

07/13/2023 07:08:25 PM

Reviewed on \_

Printed by : LUPIN

5/N 207YAXH03863



### Results

Run Date 12/07/2023 12:12:55 PM

**Last Name** First Name Gender Patient ID Birth Date

Sample

Age

Range

11.0 - 22.0

44 - 140

18.0 - 50.0

Operator ABX

Sample ID HA00342671 Rack/Pos

Department Physician

Type Standard

	comm	ents
RBC	3.29	L
HGB	9.6	1
HCT	28.7	L
MCV	87.1	

13.1

131

21.6

PDW

P-LCC

P-LCR

RBC	3.29	L	10°/µL	3.80 - 6.00
HGB	9.6	1	g/dL	11.5 - 17.0
HCT	28.7	L	%	35.0 - 52.0
MCV	87.1		µm <sup>2</sup>	76.0 - 100.0
MCH	29.0		pg	27.0 - 34.0
мснс	33.4		g/dL	32.0 - 35.0
DW-CV	20.9	Н	%	11.0 - 17.0
DW-SD	55.4	Н	μm³	37.0 - 49.0
				Range
PLT	606	H*	10³/µL	150 - 400
PCT	0.50	h*	%	0.15 - 0.40
MPV	8.3	*	μm³	8.0 - 11.0

um<sup>3</sup>


103/µL

%

\$0 100 150	2 7 X			RBC
2 7 X	2 7 X			
]   PL	PL	50 100 1	50	
				PL
		4	20	7.0

Range 103/µL 3.50 - 10.00 WBC 170.90 H\*

	#		Range	%		Range
NEU	94.63	Н*	1.60 - 7.00	71.7	*	40.0 - 73.0
LYM	20.77	H*	1.00 - 3.00	15.7	*	15.0 - 45.0
MON	8.66	H*	0.20 - 0.80	6.6	*	4.0 - 12.0
EOS	7.89	H*	0.00 - 0.50	6.0	*	0.5 - 7.0
BAS	0.00	*	0.00 - 0.15	0.0	*	0.0 - 2.0
LIC	38.95	H*	0.00 - 0.10	29,5	H*	0.0 - 1.0



# Recommended actions Slide review

Alarms

# WBC

Background Noise LYM Interference MON Interference PLT

RBC PLT Interference Susp. Pathologies Anisocytosis Thrombocytosis PLT aggregate or NRBC Leukocytosis Lymphocytosis Neutrophilia Eosinophilia Monocytosis Large Immature Cells Left shift

Atypic Lymphocytes Malaria P. falciparum 7 Malaria P. vivax ?

# Slide Review

Myeloblast	Anisocytosis
Promyelocyte	Hypochromia
Myelocyte	Polychromasia
Metamyelocyte	Poikilocytosis
Blast	Microcytosis
Target Cell	Macrocytosis
Sickle Cell	Platelet Clumps
	Promyelocyte  Myelocyte  Metamyelocyte  Blast  Target Cell

Signature: by . Reviewed on \_

## Results

Run Date 07/13/2023 06:59 09 PM

Last Name First Name Gender

Patient ID

Komal Talcale

Operator LUPIN

Sample ID | HA00342671 Rack/Pos 041731/2

Department Physician

Type Standard

Birth Date

Sample comments

Slide review

WBC

Background Noise LYM Interference MON Interference Abnormal Differentiation

PLT

**RBC PLT Interference** 

Susp. Pathologies

Anemia Anisocytosis Thrombocytosis PLT aggregate or NRBC 7 Leukocytosis Lymphocytosis Neutrophilia Eosinophilia Monocytosis

Large Immature Cells

Left shift Atypic Lymphocytes

Recommended actions


Range 10%/HL 3.23 3.80 - 6.00 RBC 11.5 - 17.0 g/dL 9.5 HGB 36 35.0 - 52.0 28.3 HCT um3 76.0 - 100.0 87.5 MCV 27.0 - 34.0 29.3 MCH pg 32.0 - 35.0 33.5 g/dl. MCHC 9% 11.0 - 17.0 21.0 RDW-CV  $\mu m_3$ 37.0 - 49.0 63.8 RDW-SD Range 101/µL 150 - 400 782 11 3 PLT 0.15 - 0.40 % 0.76 TX PCT 8.0 - 11.0  $\mu m_3$ 9.8 MPV 11.0 - 22.0 um3 PDW 18.0 10 1/µL 44 - 140 267 P-LCC 18.0 - 50.0 % 34.2 P-LCR

180 PLI

Range

3.50 - 10.00 103/µL WBC 142.87 H\*

		#		Range	%		Range
	NEU	69.86	計冰	1.60 - 7.00	69.6	*	40.0 - 73.0
-	LYM	18.52	He	1.00 - 3.00	18.5	*	15.0 - 45.0
	MON	7.79	Pol six	0.20 - 0.80	7.8	*	4.0 - 12.0
	EDS	4.16	Him	0.00 - 0.50	4.1	*	0.5 - 7.0
	BAS	0.00	*	0.00 - 0.15	0.0	**	0.0 - 2.0
	LIC	47.44	H冰	0.00 - 0.10	42.3	HX	0.0 - 1.0



Slide Review

Neutrophil Lymphocyte

Monocyte Eosinophil Basophil

Atypical Lymphocyte

Other

Myeloblast

Promyelocyte Myelocyte Metamyelocyte

Blast Target Cell Sickle Cell

Anisocytosis

Hypochromia Polychromasia

Poikilocytosis

Microcytosis Macrocytosis

Platelet Clumps

Signature: Reviewed on

07/13/2023 07:08:41 PM

Printed by : LUPIN

5/N 207YAXH03863

14 Jul 2023, 122 pm

