



# SD BIOSENSOR HEALTHCARE PVT. LTD.

Registered Office cum R&D Center : Plot No. 63, Udyog Vihar, Phase-IV, Gurugram-122015, Haryana, India  
Tel.:(+91) 124-454-0907, Toll Free No: (+91) 1800 10 23105 Website: www.sdbiosensor.co.in : Email: care@sdbiosensor.co.in  
CIN - U33110HR2012PTC047034

## In Field Calibration Report

Standard F 200
Serial No: FA20FAITG1572




Operator
Name: SD Biosensor Healthcare Pvt.Ltd.

Instrument Owner
Name: Lupin Diagnostics, Ahmednagar

Control Devices	Lot No.	Expiry Date
CAL 1	FC0121002	01-08-2023
CAL 2	FC0121002	01-08-2023
CAL 3	FC0121002	01-08-2023
C - A1C - 1	9B23BA1AC - L1	09-05-2024
C - A1C - 2	9B23BA1AC - L2	09-05-2024

Operating Conditions	
Temperature	15- 32°C

Result of Control Device Measurements			
Control Device	Measured Value		Accepted Range
	1	2	
CAL 1	Ok	Ok	N/A
CAL 2	Ok	Ok	N/A
CAL 3	Ok	Ok	N/A
9B23BA1AC - L1	5.8	5.8	4.3 % - 6.7%
9B23BA1AC - L2	9.3	9.1	7.5 %-11.5 %

Final Result		
The operator hereby confirm that the In Field Calibrations was performed in accordance to the instructions given in "Standard F 200 , In Field Calibration Procedure" .		
Date:  17-07-2023	Authorised Operator Signature  	Instrument Responsible Signature   



Manufacturing Facility India : Plot 38, Sector 4, IMT Manesar, Gurugram-122052  
Warehouse : Khasra No.33/12/1, Kherki Daula. Sector-76, Sikhopur Road, Gurugram - 122004 (Haryana)

# STANDARD™

2023/07/17

17:51

SN : FA20FAITG1572

Version : V001.063.37

Calibration Result

Operator : guest

Result : OK

Note:

F200

# Calibration

## Result

# OK

---

CAL-1	OK
CAL-2	OK
CAL-3	OK

OK

Print

2023.07.17 | 05:52:34 PM

# SD BIOSENSOR

**SD BIOSENSOR HEALTHCARE PVT. LTD.**


Registered Office cum R&D Center : Plot No. 63, Udyog Vihar, Phase-IV, Gurugram-122015, Haryana, India  
Tel.:(+91) 124-454-0907, Toll Free No: (+91) 1800 10 23105 Website: www.sdbiosensor.co.in : Email: care@sdbiosensor.co.in  
CIN - U33110HR2012PTC047034

---


## Installation Certificate for Standard F200

This is to certify that Standard F200 instrument Serial No. FA20FAITG1572 is successfully installed and commissioned at Lupin Diagnostics, Ahmednagar and the installation protocol/checklist has been successfully completed for the above instrument.

For SD Biosensor Health Care Pvt. Ltd.

  
Vikas Chawla

Product Manager



---

Manufacturing Facility India : Plot 38, Sector 4, IMT Manesar, Gurugram-122052

Warehouse : Khasra No.33/12/1, Kherki Daula, Sector-76, SIKHOPUR ROAD, Gurugram - 122004 (Haryana)



# SD BIOSENSOR

## SD BIOSENSOR HEALTHCARE PVT. LTD.

Registered Office cum R&D Center : Plot No. 63, Udyog Vihar, Phase-IV, Gurugram-122015, Haryana, India  
Tel.:(+91) 124-454-0907, Toll Free No: (+91) 1800 10 23105 Website: www.sdbiosensor.co.in : Email: care@sdbiosensor.co.in  
CIN - U33110HR2012PTC047034

### Installation Report

Customer Details	Instrument Details
------------------	--------------------

Name: Lupin Diagnostics  
Address: Maccare Hospital, Savedi, Ahmednagar

Model: Standard F 200

Contact Person(s): Dr. Sagar Kulat  
Contact No: 9960006503

Serial Number: FA20FAITG1572  
Product of: SD Biosensor  
Installation Date: 03/10/2022.

Accessories and Consumable Supplied
-------------------------------------

1. Power adapter: Yes
2. Calibration set: Yes
3. Test kit, Buffer: Yes
4. User instruction guide: Yes

Installation Checks
---------------------

1. Environment: Closed, AC
2. Power Adapter: Working
3. Temperature: 26 degrees Celsius (in limits between 15 degrees to 32 degrees Celsius)
4. UPS: Yes


Training Imparted	Assays Performed
-------------------	------------------


1. Operating Procedure
2. Maintenance
3. Dos and Don'ts
4. Calibration

Customer Remarks	Installer's Remarks
------------------	---------------------



Customer's Signature

  
Installer's Signature



Manufacturing Facility India : Plot 38, Sector 4, IMT Manesar, Gurugram-122052  
Warehouse : Khasra No.33/12/1, Kherki Daula, Sector-76, Sikhopur Road, Gurugram - 122004 (Haryana)

# SD BIOSENSOR

**SD BIOSENSOR HEALTHCARE PVT. LTD.**


Registered Office cum R&D Center : Plot No. 63, Udyog Vihar, Phase-IV, Gurugram-122015, Haryana, India  
Tel.:(+91) 124-454-0907, Toll Free No: (+91) 1800 10 23105 Website: www.sdbiosensor.co.in : Email: care@sdbiosensor.co.in  
CIN - U33110HR2012PTC047034

---


## Operations Qualification for Standard F 200

1. Verified Cal 1 Calibration (blank check) : Done
2. Verified Cal 2 Calibration (UV light source and detector check) : Done
3. Verified Cal 3 Calibration (RBG light source and three level band  
Intensity (Strong positive, weak positive and negative) check) : Done

For SD Biosensor Health Care Pvt. Ltd.

  
Vikas Chawla

Product Manager



---

Manufacturing Facility India : Plot 38, Sector 4, IMT Manesar, Gurugram-122052  
Warehouse : Khasra No.33/12/1, Kherki Daula, Sector-76, SIKHOPUR ROAD, Gurugram - 122004 (Haryana)

