

CENTRAL LAB
SANTOSH HOSPITAL GHAZIABAD
Proficiency Testing Report

S.No.	Name of Test	Method	SDI (May-23)	Remark
1.	Glucose	GOD-POD	-0.69	Satisfactory
2.	Urea	Urease UV	-1.61	Satisfactory
3.	Creatinine	Enzymatic Colorimetry	-2.79	Satisfactory
4.	T. Bilirubin	Jendrassik	-2.43	Satisfactory
5.	T. Protein	Biuret	-0.88	Satisfactory
6.	Albumin	BCG	0.29	Satisfactory
7.	Calcium	OCPC	0.30	Satisfactory
8.	Phosphorus	Molybdate UV	0.92	Satisfactory
9.	Uric Acid	Uricase	0-2.14	Satisfactory
10.	Cholesterol	CHOD-PAP	2.59	Satisfactory
11.	Triglyceride	GPO-PAP	0.26	Satisfactory
12.	HDL	Direct Method	6.36	Corrective action Taken (Result unacceptable) Next run to be monitored
13.	Sodium	ISE-Direct	1.12	Satisfactory
14.	Potassium	ISE-Direct	1.40	Satisfactory
15.	AST	UV Kinetic Without P-5-P	2.49	Satisfactory
16.	ALT	UV Kinetic Without P-5-P	-0.48	Satisfactory
17.	ALP	PNP AMP Kinetic	-5.58	Corrective action Taken (Result acceptable) – Random Error
18.	Amylase	Enzymatic Colorimetric / G7PNP Blocked	-0.37	Satisfactory
19.	Magnesium	Colorimetric	0.00	Satisfactory



Dr. Juhi Aggarwal , MD
Biochemistry

Lab Director