



## Calibration Certificate

**Customer Name: Santosh Hospital**

**Address of Customer: Ghaziabad**

**Model : CHEM 7 NS**

**Serial No. : N101041**

**Calibration Done Date : 09/11/2022**

**Calibration Due Date : On or Before 08/11/2023**

*This is to certify that we have calibrated the Semi-Automatic Biochemistry Analyzer Model: **CHEM7 NS** bearing **Sr. No. N101041** and on inspection of the instrument gains. It is observed that the results are well within the range and instrument is working fine.*

Thanking you,

For

**Transasia Bio-Medicals Ltd.**



**Lavkush Lal**

**Application Specialist**



## Calibration Data

S.N.	Filter	Gain Values	OK/Yes/No
1	340	10410	OK
2	405	16450	OK
3	450	28451	OK
4	505	30541	OK
5	546	27681	OK
6	578	25641	OK
7	600	28654	OK
8	670	34587	OK

Thanking you,

For

**Transasia Bio-Medicals Ltd.**



**Lavkush Lal**

**Application Specialist**