



TMSCC

TESTING MACHINE SERVICE AND CALIBRATION CENTRE
Precision is Our Destination.....

An NABL Accredited Calibration Laboratory
A Constituent Board of Quality Council of India

An ISO 9001 : 2015 Certified company

Calibration of Various Instruments & Testing Machines...



CC - 3125

Form No.: TMSCC/R/23

CALIBRATION CERTIFICATE

CALIBRATION CERTIFICATE NO.: 2023/06/492 Page 1 of 1

ISSUE DATE: 12-Jun-2023

ULR No.: CC312523000007375F

1.0 ISSUED TO: MSVP
College of Medicine
Sagore Dutta Hospital
578 B.T Road, Kolkata -700058



1.1 Service Request Form No.: SRF/2023/06/09/01

1.2 Service Request Date: 09-06-2023

1.3 Location: At Lab

Description identification of item to be calibrated:

1.4	i	Name:	Micro Pipette	ii	Make:	Microlit
	iii	Model / Type No.:	--	iv	S/L No.:	22108173
	v	I.D.No.:	--	vi	Job Code No:	2023/06/492
	vii	Range:	(100 to 1000) μ l	viii	Resolution:	1 μ l
	ix	Accuracy	As Per ISO 8655-6	x	End User :	--

1.5 Full / Partial Calibration: Full Calibration

1.6 Applicable specification of item to be calibrated: Accuracy / permissible limit: Not Specified.

1.7 Date of receipt of item: 9-Jun-2023 1.8 Date of calibration: 9-Jun-2023

1.9 Calibration due on: 8-Jun-2024 2.0 Frequency of calibration once in: 12 Months

2.1 Environmental condition during calibration
Temperature: 20.3 °C
Humidity: 53% RH

2.2 Basis of calibration: SOP/04/02

2.3 Traceability : Standards used for calibration are traceable to National Standards through NABL accredited Laboratory.

Name of the Instrument	Sl. No./ Id No.	Certificate No.	Lab Certificate No.	Calibrated On	Due On
Digital Weighing Balance	(TMSCC/FB/01) (Sl. No.- 14255716)	TSC/22-23/12101-25	CC - 2231	15-11-2022	15-11-2023
Digital Temperature Indicator with Sensor(RTD)	Sl. No.- 18K588073 (Sensor Sl. No.- 19102403)	TSC/22-23/12095-36		17-11-2022	17-11-2023

CALIBRATION RESULTS

Recorded Temperature : 20.3°C

Sl. No.	Denomination Volume in μ l	Observed Reading at Ref. Std. Balance in g (Avg. of five readings)	Density of De ionized Water in g/ml	Actual volume in μ l	Error in μ l	MPE in \pm μ l	Measurement Uncertainty in \pm μ l	Acceptance Criteria
1	100	0.09999	0.998183	100.17111	0.17111	0.8	1.95	Pass
2	300	0.29997	0.998183	300.51333	0.51333	4.0	1.95	Pass
3	500	0.49995	0.998183	500.85555	0.85555	4.0	1.95	Pass
4	700	0.69992	0.998183	701.19778	1.19778	8.0	1.95	Pass
5	1000	0.99989	0.998183	1001.71111	1.71111	8.0	1.95	Pass

Measurement Uncertainty at 95% confidence level where coverage factor, k=2

Remarks: The above DUC has been calibrated over its above range & the readings observed are tabulated above. The reference standard used is traceable to National Standard.

DUC: Device Under Calibration

Physical status of the Instrument : Ok

Calibrated By :

P. Dasmalakar
(Sr. Calibration Engineer)

Checked By :

J. Bhattacharjee
(Sr. Calibration Engineer)

Approved By :

C. Ghosh
(Technical Director)



END OF CERTIFICATE