



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - NOVEMBER 2022



PC-1024

Lab Name Govt. PHC, Madalapattu

Lab No 12064

Constituent Group Chemistry III

Date of Result Entered : 14/11/2022

PT Item Lyophilized Serum

Date of Report Published : 06/12/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1868	82.95	8.62	7.15	106 mg/dL	1.22	0.33
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1843	6.98	7.13	0.50	3.3 mg/dL	-1.36	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1767	112.60	5.74	6.46	169 mg/dL	0.73	0.31

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Madalapattu

THE MEDICAL OFFICER PHC MADALAPATTU

Cuddalore

TAMILNADU605007

Coordinator Contact Details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College, Vellore

S. Shylaja
The Medical officer
Govt. Primary Health Centre
Madalapattu - 607402.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

