



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - DECEMBER 2022



PC-1024

Lab Name Govt. PHC, Madalapattu  
Constituent Group Chemistry III  
PT Item Lyophilized Serum

Lab No 12054  
Date of Result Entered : 15/12/2022  
Date of Report Published : 05/01/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1844	239.20	6.25	14.96	130 mg/dL	-0.3	0.70
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1839	3.91	8.24	0.32	0.4 mg/dL	-1.97	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1766	114.94	5.41	6.21	178 mg/dL	1.15	0.30

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

Govt. PHC, Madalapattu  
THE MEDICAL OFFICER PHC MADALAPATTU  
Cuddalore  
TAMILNADU 605007

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*S. Shylaja*  
The Medical officer  
Govt. Primary Health Centre  
Madalapattu - 607402.

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential

