



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JANUARY 2023



PC-1024

Lab Name Govt. PHC, Madalapattu
Constituent Group Chemistry III
PT Item Lyophilized Serum

Lab No 12064
Date of Result Entered : 09/01/2023
Date of Report Published : 09/02/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1890	119.87	6.38	7.65	89 mg/dL	2.04	0.35
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1865	2.21	10.02	0.22	5.2 mg/dL	0.59	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1794	112.14	6.03	6.76	158 mg/dL	1.78	0.32

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Madalapattu
THE MEDICAL OFFICER PHC MADALAPATTU
Cuddalore
TAMILNADU 605007

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

S. Shylaja
The Medical officer
Govt. Primary Health Centre
Madalapattu - 607402.

