

		CA	LIBRA	TION CE	RTIFIC	ATE	
CERTIFICATE NO: SBS/CL/23/01971				MEDICAL DEVICES		VICES	Page No:1 of 1
Issue	Date		4 1	20-02-2023	5,44		
SRF No & Date			SRF/23/00067-0002 & 19-02-2023				
Receipt Date			19-02-2023				
Calibration Date			19-02-2023				
Calibration Due			18-02-2024				
Custo	mer Name & Address	12			100 pt 51 100		
GOV	RNMENT PRIMARY HE	ALTH CENTR	E,	- 1 - 1	* 15# - 1 - 2		
MATH	IALAPATTU, CUDDALORI	E-605007.				1 2 2 2 2 2 2 E	
6 48 1			Details of	Device Under Ca	alibration (DUC	C)	
Description : SEMI AUTO ANALYZ Range : Resolution :		ER	Make & Model Sr. No Identification No		ROBONIK & PRIETE:	ST TOUCH	
		FACTORY		Location		LABORATORY	
000	Sorialion . Ortifo		onmental Co	onditions & Calibi	ration Procedu	ire Details	
Environmental Details Temperature: 25.6°C			Relative Hu	ALC: 7 (4)	54% RH		
		SBS/CP/MD/2		Calibration done at		ONSITE	
		TO SE	Total Control of the	ference Standard	s Details	The same of the sa	
	S.No Description Make/		Make/ SI No	ke/ SI No:		No	154 41 414
S.No	Description	*					Validity

ELECTRICAL SAFETY

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance	Measured values in MΩ	Uncertainty in % (±)
	>20ΜΩ	96	13.92
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)
	<5000µAfor B,BF,CF	208	7.2
3	Enclosure Leakage	Measured values in μA	Uncertainty in % (±)
	<500µAfor B,BF,CF	171	8.3

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

C.SIVABALAN





Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

