

			CA	LIBR/	TIC	ON C	ER	TIF	-ICA	ATE	
CERTIFICATE NO: SBS/CL/23/01972					MEDICAL DEVICES						Page No:1 of 1
Issue Date					20-02-2023						
SRF No & Date				SRF/23/00067-0003 & 19-02-2023							
Receipt Date				19-02-2023							
Calibration Date				19-02-2023							
Calibration Due					18-02-2024						
Customer Name &	Address	_			•						
GOVERNMENT	PRIMAR'	Y HEAL	TH CENTR	E,							
MATHALAPATTU	J,CUDDA	LORE-	605007.								
				Details o	f Devic	e Under	Calibi	ation	(DUC)		
Description ELECTRICAL SAFETY(MICROSCOPE)			PE)	Make & Model : LABOMED & VISION 2000					000		
Range :			Sr. No 200220596			200220596					
Resolution :			Identification No :								
DUC Condition	: 3	SATISFA	ACTORY		Locat	ion			:	LABORATORY	
			Envi	ronmental C	onditio	ns & Cali	ibratio	n Pro	cedure	e Details	
Environmental Details Temperature:			25.6°C	Relative Humi			idity 54% RH		54% RH		
Calibration Procedure No SBS/CF			SBS/CP/MD/2	/29		Calibration done at		ONSITE			
				R	eferenc	e Standa	rds D	etails	7		
S.No Description				Make/ SI No	Make/ SI No:			Certificate No			Validity
Electrical Safety Analyser			Rigel Medical & 44L-1059				TSC/22-23/7400-3			10-08-2023	

## **ELECTRICAL SAFETY**

RESULTS						
S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)			
1	Insulation Resistance	Measured values in $M\Omega$	Uncertainty in % (±)			
	>20ΜΩ	100	13.92			
2	Earth Leakage	Measured values in μA	Uncertainty in % (±)			
	<5000µAfor B,BF,CF	123	12.1			
3	Enclosure Leakage	Measured values in μA	Uncertainty in % (±)			
	<500µAfor B,BF,CF	171	8.3			
REMARKS	3					

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00. 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.
- Calibrated By

**Authorised Signatory** 

C.SIVABALAN



Chief Executive

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

