



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name **DISTRICT HOSPITAL SHAHDOL** Lab No **7818**  
Constituent Group **HbA1c** Date of Result Entered : **12/06/2023**  
PT item **Lyophilized human whole blood based** Date of Report Published : **04/07/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	HPLC-Ion Exchange	Biorad ( D10 / D100 / Variant Turbo )	715	6.80	2.65	0.18	6.8 %	0.00	0.01

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

**DISTRICT HOSPITAL SHAHDOL  
WARD NO,05 SHRI HUSHABHAU THAKRE, DH SHAHDOL  
SHAHDOL  
MADHYA PRADESH484001**

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*