



BIO-RAD LABORATORIES (INDIA) PRIVATE LIMITED

Site name:	GOVT. DISTRICT HOSPITAL SHAHDOL
Address:	PONDOV NARAG SHAHDOL
Province-Postal code	484001
Contact Name	Mr.DEEPAAK VARMA

**Calibration Certificate**

Date	17-05-2023 <small>(dd mmmm, yyyy)</small>	Instrument Model	D-10 Hb Analyzer	Instrument Serial Number	DJ22L08805
------	--	------------------	------------------	--------------------------	------------

We certify that the Bio-Rad instrument specified above and installed in your lab is calibrated and the performance was found satisfactory.

**High Pressure Pump Flow Rate Reference: Service Manual: Section 7.7.3**

Measured with <small>(Device manufacturer and model)</small>	Device Serial Number	Device Calibration date <small>(dd mmmm, yyyy)</small>
Measuring cylinder hex base 10ML	2310100	15.05.23
Specified Range@ 1.5ml/Minute	Measured Value	
1.425 – 1.575 ml/minute	1.49	

**Column Holder temperature Reference: Service Manual: Section 7.9**

Measured with <small>(Device manufacturer and model)</small>	Device Serial Number	Device Calibration date <small>(dd mmmm, yyyy)</small>
FLUKE 179	26030078	15.05.2023
Specified Range @ 35°C	Measured Value	
33.4°C – 36.4°C	34.4	

**Detector log value Reference: Service Manual: Section 7.2 or Field Bulletin D-10-081**

Measured with	Specified Range	Measured Value
Instrument Service Software	0.018 – 0.022 AU	0.021AU

The next calibration is due on	16-05-2024	<small>(dd mmmm, yyyy)</small>
--------------------------------	------------	--------------------------------

For Bio-Rad Laboratories India Pvt Limited

Authorized Signatory: