



# ARSH ENTERPRISE

(NABL Accredited Laboratory as per ISO/IEC 17025)

## A House Of Calibration

A-337, Sumel Business Park-4, Kalupur-Naroda Road, Amdupura-382345, Ahmedabad.  
M. 9974712437, 9173769174 | E-mail : arshenterprise1992@gmail.com



### Calibration Certificate

Discipline : Electro-Technical ULR/Certificate No.: CC312623000015748F

#### 1) Customer Information:

Customer Details:	<b>District Govt. Hospital Shahdol</b>	SRF No.	AE/SRF/23/1855
	<b>Central Processing Lab.</b>	Receipt Date	12/07/2023
	Kusha Bhau Thakrey District Hospital,	Calibration Date	12/07/2023
	Pandav Nagar Shahdol MP - 484001.	Due Date	11/07/2024
		Date of Issue	12/07/2023

#### 2) Instrument Information:

Calibrated Item	<b>Temperature Controller Of Refrigerator</b>	Visual Inspection	Ok
Make/Model	Haier / ===	Type	Digital / RTD PT-100
Range	2 to 8 °C	Least Count	1 °C
Serial No./ID No.:	HVC340GHC / DH SDL/CPL/REF./24	Calibration Performed	At Site
User Location	Lab	Parameter	Temperature Simulation

#### 3) Environmental Conditions:

Temperature	25 ± 5 °C	Calibration Ref Std No.	EURAMET cg-11
Relative Humidity	50 ± 20 % RH	Calibration Procedure No.	AE/CAL/ET/SOP-03

#### 4) Reference Standards Used:

Reference Standard	Serial No.	Certificate No.	Valid Up To	Traceability
Multifunction Calibrator	994413890	AMC/00000226	13/02/2024	CC-2364

#### 5) Calibration Results:


Set Reading On Master	Reading Observed on UUC	Error (±)	Expanded Uncertainty (±°C)
2.0 °C	2 °C	0 °C	1.11
3.0 °C	3 °C	0 °C	
4.0 °C	4 °C	0 °C	
6.0 °C	6 °C	0 °C	
8.0 °C	7 °C	-1 °C	

#### 6) Notes:

1. For All Notes Kindly Refer Overleaf

Calibrated By

Anas Mansuri  
(Calibration Engineer)



Authorised Signatory

Mansuri Md.Arif  
(Quality Manager)

- End of Certificate -