



# ARSH ENTERPRISE

(NABL Accredited Laboratory as per ISO/IEC 17025)

## A House Of Calibration

A-337, Sumel Business Park-4, Kalupur-Naroda Road, Amdupura-382345, Ahmedabad.  
M. 9974712437, 9173769174 | E-mail : arshenterprise1992@gmail.com



### Calibration Certificate

Discipline : Mechanical Certificate No.: CC312623000015749F

#### 1) Customer Information:

Customer Details:	<b>District Govt. Hospital Shahdol</b>	SRF No.	AE/SRF/23/1855
	<b>Central Processing Lab.</b>	Receipt Date	12/07/2023
	Kusha Bhau Thakrey District Hospital,	Calibration Date	12/07/2023
	Pandav Nagar Shahdol MP - 484001.	Due Date	11/07/2024
		Date of Issue	12/07/2023

#### 2) Instrument Information:

Calibrated Item	<b>Centrifuge Machine</b>	Visual Inspection	Ok
Make/Model	Laby / ===	Type	Digital
Range	5000 RPM	Calibration Performed	At Site
Serial No./ID No.:	2019-20/12 / DH SDL/CPL/CENTRIFUGE/20	Parameter	Speed
User Location	Lab	Least Count	1 RPM

#### 3) Environmental Conditions:

Temperature	23 ± 5 °C	Calibration Ref Std No.	===
Relative Humidity	50 ± 20 % RH	Calibration Procedure No.	AE/CAL/M/SOP-27

#### 4) Reference Standards Used:

Reference Standard	Serial No.	Certificate No.	Valid Up To	Traceability
Digital Tachometer	AE/M/28	PE23CC02645	22/05/2024	CC-3034

#### 5) Calibration Results:


Reading Observed in UUC	Reading shown in Master Digital Tachometer	Error(±)
3000 RPM	2999.2 RPM	-0.8 RPM
3500 RPM	3498.8 RPM	-1.2 RPM
4000 RPM	3998.1 RPM	-1.9 RPM
5000 RPM	4997.9 RPM	-2.1 RPM

#### 6) Notes:

1. For All Notes Kindly Refer Overleaf

Calibrated By

Yash Khatri  
(Calibration Engineer)



Authorised Signatory

Mansuri Md.Arif  
(Quality Manager)

- End of Certificate -