



ARSH ENTERPRISE

(NABL Accredited Laboratory as per ISO/IEC 17025)

A House Of Calibration

A-337, Sumel Business Park-4, Kalupur-Naroda Road, Amdupura-382345, Ahmedabad.
M. 9974712437, 9173769174 | E-mail : arshenterprise1992@gmail.com



Calibration Certificate

Discipline : Mechanical

Certificate No.:

CC312623000015751F

1) Customer Information:

Customer Details:	District Govt. Hospital Shahdol Central Processing Lab. Kusha Bhau Thakrey District Hospital, Pandav Nagar Shahdol MP - 484001.	SRF No.	AE/SRF/23/1855
		Receipt Date	12/07/2023
		Calibration Date	12/07/2023
		Due Date	11/07/2024
		Date of Issue	12/07/2023

2) Instrument Information:

Calibrated Item	Centrifuge Machine	Visual Inspection	Ok
Make/Model	Unicef / ===	Type	Digital
Range	5000 RPM	Calibration Performed	At Site
Serial No./ID No.:	=== / DH SDL/CPL/CENTRIFUGE/30	Parameter	Speed
User Location	Lab	Least Count	1 RPM

3) Environmental Conditions:

Temperature	23 ± 5 °C	Calibration Ref Std No.	===
Relative Humidity	50 ± 20 % RH	Calibration Procedure No.	AE/CAL/M/SOP-27

4) Reference Standards Used:

Reference Standard	Serial No.	Certificate No.	Valid Up To	Traceability
Digital Tachometer	AE/M/28	PE23CC02645	22/05/2024	CC-3034

5) Calibration Results:

Reading Observed in UUC	Reading shown in Master Digital Tachometer	Error(±)
3000 RPM	3000.9 RPM	0.9 RPM
3500 RPM	3501.6 RPM	1.6 RPM
4000 RPM	4002.9 RPM	2.9 RPM
5000 RPM	5003.3 RPM	3.3 RPM

6) Notes:

1. For All Notes Kindly Refer Overleaf

Calibrated By

Yash Khatri

(Calibration Engineer)



Authorised Signatory

Mansuri Md.Arif

(Quality Manager)

- End of Certificate -