



ARSH ENTERPRISE

(NABL Accredited Laboratory as per ISO/IEC 17025)

A House Of Calibration



A-337, Sumel Business Park-4, Kalupur-Naroda Road, Amdupura-382345, Ahmedabad.
M. 9974712437, 9173769174 | E-mail : arshenterprise1992@gmail.com

Calibration Certificate

Discipline : Mechanical ULR/Certificate No.: CC312623000015746F

1) Customer Information:

Customer Details:	District Govt. Hospital Shahdol	SRF No.	AE/SRF/23/1855
	Central Processing Lab.	Receipt Date	12/07/2023
	Kusha Bhau Thakrey District Hospital,	Calibration Date	12/07/2023
	Pandav Nagar Shahdol MP - 484001.	Due Date	11/07/2024
		Date of Issue	12/07/2023

2) Instrument Information:

Calibrated Item	Micro Pipette	Visual Inspection	Ok
Make/Model	Dragon / ===	Type	===
Range	5 to 50 μ l	Least Count	0.1 μ l
Serial No/ ID NO:	YE209A50035278 / DH SDL/CPL/DRAGON/19	Calibration Performed	At Lab
User Location	Lab	Parameter	Volume

3) Environmental Conditions:

Temperature	20 \pm 2 $^{\circ}$ C	Calibration Ref Std No.	ISO 8655
Relative Humidity	50 \pm 10 % RH	Calibration Procedure No.	AE/CAL/M/SOP-18

4) Reference Standards Used:

Reference Standard	Serial No.	Certificate No.	Valid Up To	Traceability
Digital Analytical Weighing Balance	20181221	EWS-1022-082	18/10/2023	CC-3295

5) Calibration Results:

Nominal Value On UUC	Reading Observed By Master	(\pm) Error
20.0 μ l	20.14 μ l	0.14 μ l
30.0 μ l	30.21 μ l	0.21 μ l
50.0 μ l	50.35 μ l	0.35 μ l

6) Notes:

- For All Notes Kindly Refer Overleaf
- Expanded Uncertainty : \pm 150 μ l

Calibrated By

Yash Khatri
(Calibration Engineer)



Authorised Signatory

Mansuri Md.Arif
(Quality Manager)

- End of Certificate -