



ARSH ENTERPRISE

(NABL Accredited Laboratory as per ISO/IEC 17025)

A House Of Calibration

A-337, Sumel Business Park-4, Kalupur-Naroda Road, Amdupura-382345, Ahmedabad.
M. 9974712437, 9173769174 | E-mail : arshenterprise1992@gmail.com



Calibration Certificate

Discipline : Mechanical ULR/Certificate No.: CC312623000015747F

1) Customer Information:

Customer Details:	District Govt. Hospital Shahdol	SRF No.	AE/SRF/23/1855
	Central Processing Lab.	Receipt Date	12/07/2023
	Kusha Bhau Thakrey District Hospital,	Calibration Date	12/07/2023
	Pandav Nagar Shahdol MP - 484001.	Due Date	11/07/2024
		Date of Issue	12/07/2023

2) Instrument Information:

Calibrated Item	Micro Pipette	Visual Inspection	Ok
Make/Model	Dragon / ===	Type	===
Range	100 to 1000 µl	Least Count	1 µl
Serial No/ ID NO:	YEABAD0014917 / DH SDL/CPL/DRAGON/18	Calibration Performed	At Lab
User Location	Lab	Parameter	Volume

3) Environmental Conditions:

Temperature	20 ± 2 °C	Calibration Ref Std No.	ISO 8655
Relative Humidity	50 ± 10 % RH	Calibration Procedure No.	AE/CAL/M/SOP-18

4) Reference Standards Used:

Reference Standard	Serial No.	Certificate No.	Valid Up To	Traceability
Digital Analytical Weighing Balance	20181221	EWS-1022-082	18/10/2023	CC-3295

5) Calibration Results:

Nominal Value On UUC	Reading Observed By Master	(±) Error
100 µl	100.12 µl	0.12 µl
500 µl	501.23 µl	1.23 µl
1000 µl	1001.29 µl	1.29 µl

6) Notes:

- For All Notes Kindly Refer Overleaf
- Expanded Uncertainty : ± 150 µl

Calibrated By

Yash Khatri
(Calibration Engineer)



Authorised Signatory

Mansuri Md.Arif
(Quality Manager)

- End of Certificate -