

HIN/MED/2022-23/07301501

30th July, 2022

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer of **HORIBA MEDICAL: ABX PENTRA XL 80** bearing serial no: - **612PXL7649** installed at **APOLLO DIAGNOSTICS, Hyderabad** was calibrated with calibrator on **29th July, 2022**.

Calibrator : **ABX MINOCAL**
Lot No : **CX472**
Expiry Date : **05th Aug, 2022**

Coefficient	WBC	RBC	HGB	HCT	PLT
Target	9.0	4.50	13.6	37.4	238
Mean	8.89	4.47	13.54	37.09	236.01
Observed CV%	1.0	0.5	0.18	0.54	1.53
Accepted CV%	<2%	<2%	<1%	<2%	<5%
Status	Passed	Passed	Passed	Passed	Passed

The reports of Blank cycle, Repeatability and Calibration values were all found in acceptable range.

Next Calibration is due on **30th July, 2023**.



M.OMPRASKASH

ENGINEER –CS

For **Horiba India Pvt Ltd**