



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JANUARY 2023



PC-1024

Lab Name: Govt. Block PHC, Thalainayiru
Constituent Group: Chemistry III
PT Item: Lyophilized human serum based

Lab No: 7117
Date of Result Entered: 20/01/2023
Date of Report Published: 09/02/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1890	119.87	6.38	7.65	112.3 mg/dL	-0.99	0.35
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1865	2.21	10.02	0.22	2.48 mg/dL	1.22	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1794	112.14	6.03	6.76	116 mg/dL	0.57	0.32
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
Govt. Block PHC, Thalainayiru
The Block Medical Officer, Thalainayiru, Nagapattinam District
Thalainayiru
TAMILNADU614712

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christodoss
Dr. Pamela Christodoss
CMC EQAS Coordinator
Christian Medical College, Vellore
MEDICAL OFFICER,
GOVT. PRIMARY HEALTH CENTRE,
THALANAYAR-614 712.

S. M. S.
(19/01/23)

Homogeneity and Stability of the sample is passed.
Data In CMC EQAS reports is confidential
CMC EQAS does not sub contract any components