





CALIBRATION REPORT	
Report No : HBMS/CAL/22-23	
Calibration date : 12.09.2022	
Calibration valid: 11.08.2023	
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/S. GOVT.PRIMARY HEATH CENTER, THIALAINAYAR, VEDARANYAM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
I Manufactured by	VIBASH
II Model	10-100 µl
III Serial No	03435PW

Details of the machine : machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.03	50.06	100.14
2	10.04	50.09	100.12
3	10.05	50.02	100.10
4	10.09	50.08	100.08
5	10.00	49.94	100.09
6	10.08	50.07	100.12

For Heart Beat Medical Systems,


(Authorized signatory)


No. 23, 2nd Floor, jayasimmapuram,
Pappanaikenpalayam, Coimbatore – 37.

99400 28355, 97512 06106
88833 15662

hbmedicalsystem@gmail.com
hbmssales@gmail.com



CALIBRATION REPORT	
Report No: HBMS/CAL/22-23	Calibration date: 12/09/2022
	Calibration due: 11/08/2023
1.1. CUSTOMER DETAILS:	
Name and Address of the organization	M/s, GOVT.PRIMARY HEALTH CENTER, THALAINAYAR, VEDARANYAM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
I Manufactured by	THERMOSCIENTIFIC
II Model	-----
III Valuc	100µl - 1000µl
IV Serial No	RW12457

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.06	500.08	999.98
2	100.07	500.09	1000.01
3	100.01	500.05	999.96
4	90.99	499.99	1000.02
5	100.02	500.09	1000.08
6	100.01	490.94	1000.05

For Heart Beat Medical Systems,


(Authorized signatory)



No. 23, 2nd Floor, jayasimmapuram,
Pappanaikenpalayam, Coimbatore - 37.

99400 28355, 97512 06106
88833 15662

hbmedicalsystem@gmail.com
hbmssales@gmail.com