

CALIBRATION REPORT

Refrigerators, Laboratory(Lab refrigerator)



Report No:	CAL-AR-2023-02-0384	Date:12-01-2023	Page 01 of 02
------------	---------------------	-----------------	---------------

Customer Details			
Health Facility	Tezu DH		
State	Arunachal Pradesh	District	Lohit

Medical Equipment Details			
ME Number	ARZ1LOHDH0050286	Manufacturer	Whirlpool
ME Name	Refrigerators, Laboratory(Lab refrigerator)	Model	Quick-chill
Department	ICTC, Near EPI Room	Serial No	NA
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	03-01-2023	Due for Calibration	02-01-2024

Environmental Conditions			
Temperature in °C	20°C	Humidity in % Rh	58%

Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	TRI-TE-0187	DIGITAL THERMOMETER	METRAVI	DTM-900	271926/1	15-11-2023

Performance Results						
No	Parameter	Set Temp	Unit	Measured Value	Acceptable Range	Remarks
1	Temperature	2°C	°C	2.1	1.5 to 2.5°C	Pass
2	Temperature	4°C	°C	4.3	3.5 to 4.5°C	Pass

14

CALIBRATION REPORT 1

Refrigerators, Laboratory(Lab refrigerator)



3	Temperature	6°C	°C	6.2	5.5 to 6.5°C	Pass
---	-------------	-----	----	-----	--------------	------

Conditions of Issue of report

1. The Standards used are traceable to National standards.
2. This report should not be reproduced except in full without written approval from Trimed Solutions India Private Limited.

Calibration Status

Passed

Failed

Calibrated By

Surojit Dey

Authorised Signatory

Rahul Kumar Kashyap